

IDENTIFYN RESOLUTION STEPDOWN

SECONDARY ANTIBODY MICROSCOPY EVIDENCE ASSESSMENT

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse, F(ab')₂ (H + L)

WESTERN BLOT -> WIDEFIELD -> WIDEFIELD SIM / APOTOME -> CONFOCAL -> AIRYSCAN
SIM -> SIM²

7


ORDERED EVIDENCE TIERS

1

SUPPORTING CONTROL

PENDING

SCIENTIST REVIEW



REPRESENTATIVE HIGHEST-AVAILABLE VIEW / SIM²

SA-000051 | PRODUCT-LEVEL EVIDENCE DOSSIER

SOURCE-GROUNDED MICROSCOPY + BIOLOGICAL REVIEW / IDENTIFYN EVIDENCE PRESERVED

CRO CAPABILITY DEMONSTRATED

Identifyn performed the experimental and imaging work documented below. BENNETT converts that operating history into a governed, source-bounded scientific record.

- Source-documented sample preparation and antibody/reagent workflow
- Western blot imaging
- Widefield fluorescence microscopy
- Widefield structured illumination / Apotome optical sectioning
- Confocal microscopy
- Airyscan acquisition and reconstruction workflow
- Structured illumination microscopy
- SIM² reconstruction workflow
- Preserved control experiment
- Z-stack / 3D acquisition
- Multi-channel imaging

SCIENTIFIC QUESTION

What source-bounded experimental and microscopy evidence is preserved for Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse, F(ab')₂ (H + L), and what is the strongest interpretation allowed by the available modalities and controls?

EVIDENCE AVAILABLE

Western blot -> Widefield -> Widefield SIM — Apotome -> Confocal -> Airyscan -> SIM -> SIM² -> Control

BENNETT ASSESSMENT

The preserved SA-000051 record contains Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². Missing modalities are not represented or inferred.

BENNETT / MICROSCOPY AND BIOLOGICAL EVIDENCE REVIEW

SOURCE-GROUNDED RETROSPECTIVE REVIEW / PENDING SCIENTIST REVIEW

SOURCE STATE	SOURCE HASH VERIFIED / EVENT RECORDED
BENNETT EVIDENCE REVIEW	COMPLETED / SOURCE-BOUNDED
NATIVE IMAGE REPROCESSING	NOT PERFORMED
LITERATURE SOURCES	INDEPENDENTLY VERIFIED
SCIENTIST REVIEW	PENDING
PUBLICATION	NOT AUTHORIZED

OVERALL CALL / PARTIAL STEPDOWN / SOURCE-BOUNDED REVIEW

The preserved SA-000051 record contains Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². Missing modalities are not represented or inferred.

SETTINGS ASSESSMENT

The record preserves 7 ordered evidence tier(s). Structured source metadata values are reported by stage and remain tied to the original method objects.

BIOLOGICAL SUPPORT

The record supports retrospective, source-bounded microscopy review. Target identity, specificity, treatment effect, binding, interaction and mechanism are not inferred from presentation images alone.

CONTROL ASSESSMENT

A true preserved control record is present and appears last.

CLAIM CEILING

This draft supports historical capability documentation, source registration, modality ordering, settings review and bounded interpretation. It does not independently validate antibody specificity, quantify a population response, establish binding, interaction, causation or mechanism, or authorize publication.

PROCESSING DISCLOSURE

Retrospective BENNETT architectural evaluation of preserved Identifyn source evidence. BENNETT did not perform native-file reprocessing, new deterministic measurement or the historical image post-processing represented in this record.

REAGENTS AND ACQUIRED CHANNELS

Preparation reagents are reported separately from channels actually documented in the acquisition settings. Fluorophores mentioned in staining are not automatically treated as acquired channels.

EVIDENCE OBJECT	REAGENTS / FLUOROPHORES PRESENT	CHANNELS ACTUALLY ACQUIRED
Western blot	594	Not deterministically stated in the acquisition block
Widefield	405/DAPI; 594	2; 45 Texas Red; 49 DAPI; 540 - 580; 335 - 383; 593 - 668; 420 - 470; 590
Widefield SIM — Apotome	405/DAPI; 594	2; 45 Texas Red; 49 DAPI; 540 - 580; 335-383; 593 - 668; 420-470; 590
Confocal	405/DAPI; 488; 594	2; None; 561nm @1.50%; 405nm @0.3%; 590; 353; 618; 465
Airyscan	405/DAPI; 488; 594	2; None; 561nm @1.00%; 405nm @1.0%; 590; 353; 618; 465
SIM	405/DAPI; 488; 594	2; SR HE LBF 405/488/561/642; 200 - 405, 200 - 488, 200 - 561, 200 - 642; 419 - 470, 503 - 546, 576 - 617, 657 - 800; T1-Axiocam 820 #2-SR; T3-Axiocam 820 #2-SR; 561; 405
SIM ²	405/DAPI; 488; 594	2; 561nm @3.5%; 405nm @4.0%; 561; 405; 515; 561.000000; 405.000000
Control	405/DAPI; 594	2; 45 Texas Red; 49 DAPI; 540 - 580; 335 - 383; 593 - 668; 420 - 470; 590

MODALITY ASSESSMENT / PART 1 OF 8

Western blot / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: Azure Sapphire FL. Objective: not explicitly stated. Platform: Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s). Structured source metadata values: 11.

VISIBLE OBSERVATION

A preserved presentation image is available for Western blot. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Western blot, documents platform context as Azure Sapphire FL, and records detected dye/channel descriptors as 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Band position and displayed intensity do not independently establish analytical specificity, target identity, linearity, or treatment effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Western blot evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 2 OF 8

Widefield / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Imager.M1. Objective: Plan-Apochromat 63x/1.40 Oil M27. Platform: Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, AxioCam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Channels: 2; Scaling (Per Pixel): 0.102µm X 0.102µm; Image size (Pixels): 579 X 768; Image Size (Pixels): 579 X 768; Bit Depth: 12 Bit. Detector/camera: Imaging Device: AxioCam MR R3. Timing: Exposure Time: 50ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @15%. Optical/scan: Effective NA: 1.4. Structured source metadata values: 28.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield, documents platform context as ZEISS Axio Imager.M1, and records detected dye/channel descriptors as 405/DAPI; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

A presentation image supports gross localization and source-bounded co-occurrence, not direct binding, molecular colocalization, or native quantitative measurement.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 3 OF 8

Widefield SIM — Apotome / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:. Acquisition: Z-Stack: 4Slices (0.04µm); Channels: 2; Scaling (Per Pixel): 0.143µm X 0.143µm X 0.200µm; Image size (Pixels): 361 X 380; Image Size (Pixels): 361 X 380; Bit Depth: 14 Bit. Detector/camera: Imaging Device: Axiocam 807. Timing: Exposure Time: 100ms; 18ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @30%; X-Cite Xylis Lamp Intensity @6%. Optical/scan: Effective NA: 1.25. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 38.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield SIM — Apotome. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield SIM — Apotome, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Improved optical sectioning does not by itself establish reconstructed super-resolution or molecular interaction.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield SIM — Apotome evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 4 OF 8

Confocal / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 65Slices (4.48µm); Channels: 2; Scaling (Per Pixel): 70.60nm X 70.60nm X 70.00nm; Image size (Pixels): 989 X 821; Image Size (Pixels): 989 X 821; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Airyscan; GaAsP-PMT; Detector Type: GaAsP-PMT. Timing: Pixel Time: 0.70µs; Frame Time: 4.02s. Illumination: Laser Wavelength: 561nm @1.50%; 405nm @0.3%. Optical/scan: Pinhole: 1.00AU / 51µm; 1.00AU / 37µm; Scan Zoom: 1.3; LSM Scan Speed: 8; Scan Direction: Bidirectional; Averaging: 4; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Dark Grey. Structured source metadata values: 46.

VISIBLE OBSERVATION

A preserved presentation image is available for Confocal. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Confocal, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Optical sectioning and source-bounded 3D presentation do not establish binding, mechanism, or population-level effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Confocal evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 5 OF 8

Airyscan / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 136Slices (5.40µm); 136 Slices (5.40µm); Channels: 2; Scaling (Per Pixel): 35.29nm X 35.29nm X 40.00nm; Image size (Pixels): 1772 X 1772; 146 X 139; Image Size (Pixels): 1772 X 1772; 146 X 139; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Airyscan; Detector Type: GaAsP-PMT. Timing: Pixel Time: 0.59µs; Frame Time: 4.44s. Illumination: Laser Wavelength: 561nm @1.00%; 405nm @1.0%. Optical/scan: Pinhole: 5.00 AU / 251µm; 5.00 AU / 188µm; Scan Zoom: 1.6; LSM Scan Speed: 7; Scan Direction: Bidirectional; Averaging: 2; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black; AiryScan Mode: Super Resolution 9.0 (3D, Auto); Super Resolution 8.5 (3D, Auto). Structured source metadata values: 53.

VISIBLE OBSERVATION

A preserved presentation image is available for Airyscan. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Airyscan, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Displayed feature separation remains reconstruction dependent; no unsupported numerical resolution or molecular architecture claim is permitted.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Airyscan evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 6 OF 8

SIM / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:. Acquisition: Z-Stack: 91 Slices (4.50µm); Channels: 2; Scaling (Per Pixel): 16.89nm X 16.89nm X 50.00nm; Image size (Pixels): 5056 X 5056; Image Size (Pixels): 5056 X 5056; Bit Depth: 16 Bit. Detector/camera: Imaging Device: AxioCam 820 #2. Timing: Exposure Time: 120ms; 50ms; Frame Time: 1.33s; 676.00ms. Illumination: Laser Wavelength: 561nm @2.00%; 405nm @7.0%; Illumination Wavelength: 561; 405. Optical/scan: Scan Direction: Unidirectional; Effective NA: 1.4. Processing: Algorithm: SIM; Process Sampling: N/a; Result Sampling: 4; Projection: View Angle 51°, Scale Z 20%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 50.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM, documents platform context as ZEISS Elyra, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Reconstruction-dependent structural separation requires native reconstruction QC before numerical resolution or molecular dimensions can be claimed.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 7 OF 8

SIM² / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra 7. Objective: Plan-Apochromat 63X/1.46 oil Korr M27. Platform: Zeiss Elyra 7, Andor1, AxioObserver, using Plan-Apochromat 63X/1.46 oil Korr M27 was used with the following parameters for each dye: Acquisition: Z-Stack: 28 Slices (1.62µm); Channels: 2; Scaling (Per Pixel): 15.65nm X 15.65nm X 60.00nm; Image size (Pixels): 3979 X 4207; Image Size (Pixels): 3979 X 4207; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Edge 4.2 SIM; Detector Type: Camera. Timing: Exposure Time: 100ms; 50ms; Frame Time: 1.0s. Illumination: Laser Wavelength: 561nm @3.5%; 405nm @4.0%. Optical/scan: Scan Zoom: 1.0; Scan Direction: Unidirectional; Averaging: 8; Effective NA: 1.46. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 57.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM². BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM², documents platform context as ZEISS Elyra 7, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Further reconstruction-dependent segmentation does not establish molecular dimensions, structure identity, or mechanism without independent support.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM² evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 8 OF 8

Control / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Channels: 2; Scaling (Per Pixel): 0.143µm X 0.143µm; Image size (Pixels): 1183 X 872; Image Size (Pixels): 1183 X 872; Bit Depth: 14 Bit. Detector/camera: Imaging Device: AxioCam 807 mono. Timing: Exposure Time: 200ms; 30ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @25%; X-Cite Xylis Lamp Intensity @10%. Optical/scan: Effective NA: 1.25. Structured source metadata values: 32.

VISIBLE OBSERVATION

A preserved presentation image is available for Control. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Control, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed antibody-control guidance that a no-primary or related control bounds background but is not a complete specificity system.

INTERPRETIVE LIMIT

A control bounds the recorded comparison only; it does not establish universal specificity, zero background, zero cross-reactivity, or absence of free dye.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Control evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

CROSS-MODALITY SYNTHESIS

The preserved SA-000051 record contains Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². Missing modalities are not represented or inferred.

Different acquisitions are not treated as a quantitative intensity ladder. Any continuity statement remains qualitative and source bounded.

BENNETT LITERATURE CORRELATION — SOURCES INDEPENDENTLY VERIFIED

BENNETT uses independently verified primary scientific sources to test whether the interpretation is scientifically consistent and properly bounded. Literature correlation does not retroactively validate the Identifyn dataset or replace scientist review.

- Antibody and microscopy evidence require orthogonal controls and modality-appropriate interpretation; presentation images alone do not establish analytical specificity.
- Advanced-resolution presentation depends on acquisition, reconstruction or localization quality control before numerical resolution, molecular dimensions or cluster identity can be claimed.

REFERENCES

1. Uhlen M, et al. A proposal for validation of antibodies. Nat Methods. 2016;13:823-827. doi:10.1038/nmeth.3995.
2. Gustafsson MGL. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. J Microsc. 2000;198:82-87. doi:10.1046/j.1365-2818.2000.00710.x.
3. Sage D, et al. Quantitative evaluation of software packages for single-molecule localization microscopy. Nat Methods. 2015;12:717-724. doi:10.1038/nmeth.3442.

RELEASE BOUNDARY

This assessment remains a draft pending scientist review. No publication is authorized. Any future native-file analysis must enter BENNETT as a new evidence snapshot with routed engine authority, native-data QA, methods, limitations, and an independently reviewed claim ceiling.

SOURCE RECONCILIATION / QA / PROVENANCE

SOURCE STATUS - PRESERVED / AUTO-RECONCILED

Historical Identifyn source material was subjected to BENNETT retrospective QA. Original source state and correction history remain preserved in provenance.

HIGH CONFIDENCE - AUTO-RECONCILED

Confocal / Scaling (Per Pixel): 70.60nm X 70.60nm X 70.00nm -> 0.07060 μ m X 0.07060 μ m X 0.07000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=989 X 821; Image Size (Scaled)=69.82 μ m X 57.96 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.00%.

HIGH CONFIDENCE - AUTO-RECONCILED

Airyscan / Scaling (Per Pixel): 35.29nm X 35.29nm X 40.00nm -> 0.03529 μ m X 0.03529 μ m X 0.04000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=1772 X 1772; Image Size (Scaled)=62.54 μ m X 62.54 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.01%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM / Scaling (Per Pixel): 16.89nm X 16.89nm X 50.00nm -> 0.01689 μ m X 0.01689 μ m X 0.05000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=5056 X 5056; Image Size (Scaled)=85.40 μ m X 85.40 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.00%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM² / Scaling (Per Pixel): 15.65nm X 15.65nm X 60.00nm -> 0.01565 μ m X 0.01565 μ m X 0.06000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=3979 X 4207; Image Size (Scaled)=62.68 μ m X 65.85 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.65%.

BENNETT EVIDENCE REVIEW COMPLETE

INTERPRETATION TO EVIDENCE

BENNETT has completed the bounded retrospective assessment.
The following pages switch ownership to the preserved Identifyn record.



IDENTIFYN PRESERVED SOURCE EVIDENCE

ORIGINAL RECORD BEGINS

Images, source methods, acquisition settings and provenance follow in the locked
Stepdown order. Scientific wording is preserved; missing modalities are not invented.
A preserved control follows all available Stepdown tiers as supporting evidence.

PRODUCT-LEVEL STEPDOWN MAP


This dossier preserves the complete available evidence progression for SA-000051. Each tier remains tied to its original image, source method, instrument context and cryptographic identity. A preserved control is retained as supporting evidence and appears after every resolution tier.

CANONICAL STEPDOWN	EVIDENCE STAGE	INSTRUMENT	STATUS
1	Western blot	Azure Sapphire FL	PRESERVED
2	Widefield	ZEISS Axio Imager.M1	PRESERVED
3	Widefield SIM - Apotome	ZEISS Axio Observer.Z1/7 Apotome	PRESERVED
4	Confocal	ZEISS LSM 900	PRESERVED
5	Airyscan	ZEISS LSM 900	PRESERVED
6	SIM	ZEISS Elyra	PRESERVED
7	SIM ²	ZEISS Elyra 7	PRESERVED

INTERPRETIVE BOUNDARY

Ordering evidence by resolution tier does not make the modalities interchangeable. Each stage retains its acquisition environment and scientific limitations. This pilot organizes the record; it does not synthesize new biological claims.

ORIGINAL IMAGE EVIDENCE / WESTERN BLOT



EVIDENCE TIER	Western blot
INSTRUMENT	Azure Sapphire FL
CELL MODEL	HeLa
DETECTED DYES	594
METADATA VALUES	11

STRUCTURED ACQUISITION METADATA / WESTERN BLOT

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Scanner	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Detection mode	Fluorescence
Laser	532nm
Emission Filter	624±40nm
Detector	Avalanche Photodiode
Intensity	4
Pixel size	100 µm
Scan Speed	High
Quality	High
Focus	Membrane (0 µm in Z-axis)
Image	Identifyn® LLC

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

HeLa cells were homogenized using an ultrasonication probe and centrifuged to pellet down the cell debris. The lysate supernatant was mixed with NuPAGE™ LDS Sample Buffer (4X), heated at 80oC for 10 min, and subjected to electrophoresis using NuPAGE™ 4 to 12%, Bis-Tris, 1.0-1.5 mm, Mini Protein Gels in a Mini Gel Tank. The Thermo Fisher's PageRuler™ Prestained Protein Ladder, 10 to 180 kDa, was used as a molecular weight marker.

Post electrophoresis, the proteins were transferred to a polyvinylidene difluoride (PVDF) membrane using iBlot™ 3 Starter Kit, PVDF. The transfer membrane was blocked with Blocker™ Casein for 1 hour and washed 5X with PBS. The blocked membrane was incubated with Actin Monoclonal Antibody (mAbGEa) (MA1-744) for 2 hours, followed by 5X PBS washes. Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat anti-Mouse F(ab)₂ was incubated for 1 hour and washed 5X with PBS. The membrane was dried for 10 minutes at room temperature before imaging.

Scanner = Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

Scan Settings

Detection mode = Fluorescence

Laser = 532nm

Emission Filter = 624±40nm

Detector = Avalanche Photodiode

Intensity = 4

Pixel size = 100 µm

Scan Speed = High

Quality = High

Focus = Membrane (0 µm in Z-axis)

Post Processing

None

Image: Identifyn® LLC

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000051
Stage	Western blot
Instrument	Azure Sapphire FL
Microscope configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

SOURCE PROVENANCE

Structured source metadata values	11
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	3E509372698065EFD7C424EEBA4D557BF9DF3B61F224C1C4F175C0E155F29653
Method PDF SHA-256	FFC90571203F10B97848B38C97A352E8ACE27212DE4EA8651BBA53F6F5DFF194

ORIGINAL IMAGE EVIDENCE / WIDEFIELD



EVIDENCE TIER	Widefield
INSTRUMENT	ZEISS Axio Imager.M1
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 594
METADATA VALUES	28

STRUCTURED ACQUISITION METADATA / WIDEFIELD

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	Plan-Apochromat 63x/1.40 Oil M27
Channels	2
Scaling (Per Pixel)	0.102µm X 0.102µm
Image size (Pixels)	579 X 768
Image Size (Scaled)	59.28µm X 70.54µm
Bit Depth	12 Bit
Image Center Position	X: 0.00µm, Y: 0.00µm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	45 Texas Red 49 DAPI
Beam Splitter	585 395
Filter Excitation Wavelength	540 - 580 335 - 383
Filter Emission Wavelength	593 - 668 420 - 470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @15%
Exposure Time	50ms
Excitation Wavelength	590 353
Emission Wavelength	618 465
Effective NA	1.4
Imaging Device	Axiocam MR R3
Camera Adapter	1X
Depth of Focus	0.96µm 0.72µm
Binning Mode	1,1

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Nuclear Pore Complex Monoclonal Antibody (39C7), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.102µm X 0.102µm

Image size (Pixels) = 579 X 768

Image Size (Scaled) = 59.28µm X 70.54µm

Bit Depth = 12 Bit

Image Center Position = X: 0.00µm, Y: 0.00µm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

594 -

Reflector = 45 Texas Red

Beam Splitter = 585

Filter Excitation Wavelength = 540 - 580

Filter Emission Wavelength = 593 - 668

Contrast Method = Fluorescence

Light source = X-Cite Xylis Lamp Intensity @15%

Exposure Time = 50ms

Excitation Wavelength = 590

Emission Wavelength = 618

Effective NA = 1.4

Imaging Device = Axiocam MR R3

Camera Adapter = 1X

Depth of Focus = 0.96µm

Binning Mode = 1,1

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335 - 383
 Filter Emission Wavelength = 420 - 470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @15%
 Exposure Time = 50ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Imaging Device = Axiocam MR R3
 Camera Adapter = 1X
 Depth of Focus = 0.72µm
 Binning Mode = 1,1

Post Processing

None

Image Correction


None

Image by E. F. Simon
 Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000051
Stage	Widefield
Instrument	ZEISS Axio Imager.M1
Microscope configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	28
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	B83EAD159AD28E2EA26BFC0DBD92CFE79A9EB9E03332C296E4F900E75C1FF7C1
Method PDF SHA-256	6B546B82FC9BE0FAF214AA733D9C036A45CDE3AED79C017F90AE3258D95285DB

ORIGINAL IMAGE EVIDENCE / WIDEFIELD SIM - APOTOME



EVIDENCE TIER	Widefield SIM - Apotome
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 594
METADATA VALUES	38

STRUCTURED ACQUISITION METADATA / WIDEFIELD SIM - APOTOME

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Z-Stack	4Slices (0.04µm)
Channels	2
Scaling (Per Pixel)	0.143µm X 0.143µm X 0.200µm
Image Size (Pixels)	361 X 380
Image Size (Scaled)	51.57µm X 54.29µm
Bit Depth	14 Bit
Image Center Position	X: 41.02mm, Y: 48.41mm
Stage Position	X: 41.02mm, Y: 48.41mm
ROI Center Offset	X: 1.14, Y: 0.0µm
Reflector	45 Texas Red 49 DAPI
Beam Splitter	585 395
Filter Excitation Wavelength	540 - 580 335-383
Filter Emission Wavelength	593 - 668 420-470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @30% X-Cite Xylis Lamp Intensity @6%
Exposure Time	100ms 18ms
Excitation Wavelength	590 353
Emission Wavelength	618 465
Effective NA	1.25
Imaging Device	Axiocam 807
Camera Adapter	1X
Depth of Focus	1.20µm 1.00µm
Binning Mode	2,2
Apotome SIM Image Processing	Display Mode "Optical Sectioning" - enabled correction selected.
3D Maximum Projection	Mid (50% Precise - Fast)
Appearance	Threshold 13% (594), Threshold 13% (DAPI) Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 100%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Nuclear Pore Complex Monoclonal Antibody (39C7), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 4Slices (0.04µm)

Channels = 2

Scaling (Per Pixel) = 0.143µm X 0.143µm X 0.200µm

Image Size (Pixels) = 361 X 380

Image Size (Scaled) = 51.57µm X 54.29µm

Bit Depth = 14 Bit

Image Center Position = X: 41.02mm, Y: 48.41mm

Stage Position = X: 41.02mm, Y: 48.41mm

ROI Center Offset = X: 1.14, Y: 0.0µm

594 -

Reflector = 45 Texas Red

Beam Splitter = 585

Filter Excitation Wavelength = 540 - 580

Filter Emission Wavelength = 593 - 668

Contrast Method = Fluorescence

Light source = X-Cite Xylis Lamp Intensity @30%

Exposure Time = 100ms

Excitation Wavelength = 590

Emission Wavelength = 618

Effective NA = 1.25

Imaging Device = Axiocam 807

Camera Adapter = 1X

Depth of Focus = 1.20µm

Binning Mode = 2,2

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335-383
 Filter Emission Wavelength = 420-470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @6%
 Exposure Time = 18ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.25
 Imaging Device = Axiocam 807
 Camera Adapter = 1X
 Depth of Focus = 1.00µm
 Binning Mode = 2,2

Post Processing - SIM Processing

Apotome SIM Image Processing = Display Mode "Optical Sectioning" - enabled correction selected.

Normalization, Phase Correction, Fourier Filter, and Deconvolution were NOT used.

3D Image Processing

3D Maximum Projection = Mid (50% Precise - Fast)
 Appearance = Threshold 13% (594), Threshold 13% (DAPI) Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 100%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None

Image by E.F. Simon

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000051
Stage	Widefield SIM - Apotome
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	38
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT

SOURCE PROVENANCE

Cell model	HeLa
Image SHA-256	8D6FA9ECBBE94E97B991245E44A910BC8B8C293A0B8EC4F83DD2A60C160A5922
Method PDF SHA-256	CCEB4B22260D3D8ED12B853FBDB8DA6F99C622ED6C08EC8404A5054BA1C8175E

ORIGINAL IMAGE EVIDENCE / CONFOCAL



EVIDENCE TIER	Confocal
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	46

STRUCTURED ACQUISITION METADATA / CONFOCAL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	65Slices (4.48µm)
Channels	2
Scaling (Per Pixel)	0.07060 µm X 0.07060 µm X 0.07000 µm
Image size (Pixels)	989 X 821
Image Size (Scaled)	69.82µm X 57.96µm
Bit Depth	16 Bit
Image Center Position	X: 76.71mm, Y: 46.75mm
Stage Position	X: 76.71mm, Y: 46.75mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Contrast Method	Fluorescence
Pinhole	1.00AU / 51µm 1.00AU / 37µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	561nm @1.50% 405nm @0.3%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	1.3
Rotation	0°
Sampling	1.00
Pixel Time	0.70µs
Frame Time	4.02s
LSM Scan Speed	8
Scan Direction	Bidirectional
Line Step	1
Averaging	4
Excitation Wavelength	590 353
Emission Wavelength	618 465
Effective NA	1.4
Detection Wavelength	580 - 640 410 - 470
Imaging Device	Airyscan GaAsp-PMT
Detector Type	GaAsp-PMT

FIELD	SOURCE-DOCUMENTED VALUE
Detector Gain	700 V 650 V
Detector Offset	0
Detector Digital Gain	1.0
3D Volume Projection	Precise
Appearance	Threshold 6% (594), Threshold 5% (DAPI), Ramp 94% (594), Ramp 95% (DAPI)% Maximum 100% all channels
Background	Dark Grey
Light	Brightness 113%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Mitochondria Antibody (MA5-12017), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 65Slices (4.48µm)

Channels = 2

Scaling (Per Pixel) = 70.60nm X 70.60nm X 70.00nm

Image size (Pixels) = 989 X 821

Image Size (Scaled) = 69.82µm X 57.96µm

Bit Depth = 16 Bit

Image Center Position = X: 76.71mm, Y: 46.75mm

Stage Position = X: 76.71mm, Y: 46.75mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

594 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00AU / 51µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 561nm @1.50%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.3
 Rotation = 0°
 Sampling = 1.00
 Pixel Time = 0.70µs
 Frame Time = 4.02s
 LSM Scan Speed = 8
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 4
 Excitation Wavelength = 590
 Emission Wavelength = 618
 Effective NA = 1.4
 Detection Wavelength = 580 - 640
 Imaging Device = Airyscan
 Detector Type = GaAsp-PMT
 Detector Gain = 700 V
 Detector Offset = 0
 Detector Digital Gain = 1.0

DAPI -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00AU / 37µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 405nm @0.3%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.3
 Rotation = 0°
 Sampling = 1.00
 Pixel Time = 0.70µs
 Frame Time = 4.02s
 LSM Scan Speed = 8
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 4
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Detection Wavelength = 410 - 470
 Imaging Device = GaAsP-PMT
 Detector Type = GaAsP-PMT
 Detector Gain = 650 V
 Detector Offset = 0
 Detector Digital Gain = 1.0

3D Image Processing

3D Volume Projection = Precise
 Appearance = Threshold 6% (594), Threshold 5% (DAPI), Ramp 94% (594), Ramp 95% (DAPI)% Maximum 100% all channels
 Background = Dark Grey
 Light = Brightness 113%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

The background color was changed in Zen Software from the default color, black, to grey.

Image by J.K. Bennett

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000051
---------	-----------

SOURCE PROVENANCE

Stage	Confocal
Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	46
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	493A6E1779F7CECFD8987AAF2E8127E4489C02BD3A7CCBD0B340C0FB6E360265
Method PDF SHA-256	8188EB9B7D48F9805B9A8B8FA5E4CB4D0393BD413716B2B1173C8ED19FC165F7

ORIGINAL IMAGE EVIDENCE / AIRYSCAN



EVIDENCE TIER	Airyscan
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	53

STRUCTURED ACQUISITION METADATA / AIRYSCAN

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	136Slices (5.40µm) 136 Slices (5.40µm)
Channels	2
Scaling (Per Pixel)	0.03529 µm X 0.03529 µm X 0.04000 µm
Image Size (Pixels)	1772 X 1772 146 X 139
Image Size (Scaled)	62.54µm X 62.54µm 5.15µm X 4.91µm
Bit Depth	16 Bit
Image Center Position	X: 75.58mm, Y: 46.47mm
Stage Position	X: 75.58mm, Y: 46.47mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Contrast Method	Fluorescence
Pinhole	5.00 AU / 251µm 5.00 AU / 188µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	561nm @1.00% 405nm @1.0%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	1.6
Rotation	0°
Sampling	2.00
Pixel Time	0.59µs
Frame Time	4.44s
LSM Scan Speed	7
Scan Direction	Bidirectional
Line Step	1
Averaging	2
Excitation Wavelength	590 353
Emission Wavelength	618 465
Effective NA	1.4
Detection Wavelength	560-640 400-495
Imaging Device	Airyscan
Detector Type	GaAsp-PMT

FIELD	SOURCE-DOCUMENTED VALUE
Detector Gain	850 V 800 V
Detector Offset	0
Detector Digital Gain	1.0
AiryScan Mode	Super Resolution 9.0 (3D, Auto) Super Resolution 8.5 (3D, Auto)
3D Maximum Projection	Precise
Appearance	Threshold 6% (594), Threshold 5% (DAPI), Ramp 0% all channels, Maximum 100% all channels Threshold 2% (594), Threshold 2% (DAPI), Ramp 98% (594), Ramp 98% (DAPI), Maximum 100% all channels
Background	Black
Light	Brightness 100% Brightness 100%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)
3D Volume Projection	Precise

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Mitochondria Antibody (MA5-12017), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D) - Main Image

Z-Stack = 136Slices (5.40µm)

Channels = 2

Scaling (Per Pixel) = 35.29nm X 35.29nm X 40.00nm

Image Size (Pixels) = 1772 X 1772

Image Size (Scaled) = 62.54µm X 62.54µm

Bit Depth = 16 Bit

Image Center Position = X: 75.58mm, Y: 46.47mm

Stage Position = X: 75.58mm, Y: 46.47mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

Z Stack - (3D) - Image Inset

Z-Stack = 136 Slices (5.40µm)

Channels = 2

Scaling (Per Pixel) = 35.29nm X 35.29nm X 40.00nm

Image Size (Pixels) = 146 X 139

Image Size (Scaled) = 5.15µm X 4.91µm

Bit Depth = 16 Bit

Image Center Position = X: 75.58mm, Y: 46.47mm

Stage Position = X: 75.58mm, Y: 46.47mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

594 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 5.00 AU / 251µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 561nm @1.00%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.6
 Rotation = 0°
 Sampling = 2.00
 Pixel Time = 0.59µs
 Frame Time = 4.44s
 LSM Scan Speed = 7
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 2
 Excitation Wavelength = 590
 Emission Wavelength = 618
 Effective NA = 1.4
 Detection Wavelength = 560-640
 Imaging Device = Airyscan
 Detector Type = GaAsP-PMT
 Detector Gain = 850 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 AiryScan Mode = Super Resolution 9.0 (3D, Auto)

DAPI -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 5.00 AU / 188µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 405nm @1.0%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.6
 Rotation = 0°
 Sampling = 2.00
 Pixel Time = 0.59µs
 Frame Time = 4.44s
 LSM Scan Speed = 7
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 2
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Detection Wavelength = 400-495
 Imaging Device = Airyscan
 Detector Type = GaAsP-PMT
 Detector Gain = 800 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 AiryScan Mode = Super Resolution 8.5 (3D, Auto)

Post Processing - AiryScan 3D Processing

Display Mode "SR"
 Super Resolution Auto Filter Selected
 Process 3D Current Image
 Created

3D Image Processing - Main Image

3D Maximum Projection = Precise
 Appearance = Threshold 6% (594), Threshold 5% (DAPI), Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 100%
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

3D Image Processing - Inset Image

3D Volume Projection = Precise

Appearance = Threshold 2% (594), Threshold 2% (DAPI), Ramp 98% (594), Ramp 98% (DAPI), Maximum 100% all channels

Background = Black

Light = Brightness 100%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping

Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections -

None


Image by J.K. Bennett

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000051
Stage	Airyscan
Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	53
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	B86C318BA4BDEBA5A8C80FC2066075B59BE23572ADF9929A47AAAC8360BFF360
Method PDF SHA-256	C7CC73DBDC22D494D06B2F3194A783C389CAE3CCC741C7252831EF0BFD4531F9

ORIGINAL IMAGE EVIDENCE / SIM



EVIDENCE TIER	SIM
INSTRUMENT	ZEISS Elyra
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	50

STRUCTURED ACQUISITION METADATA / SIM

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	91 Slices (4.50µm)
Channels	2
Scaling (Per Pixel)	0.01689 µm X 0.01689 µm X 0.05000 µm
Image Size (Pixels)	5056 X 5056
Image Size (Scaled)	85.40µm X 85.40µm
Bit Depth	16 Bit
Image Center Position	X: 58.12mm, Y: 32.33mm
Stage Position	X: 58.12mm, Y: 32.33mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflectors	SR HE LBF 405/488/561/642
Beam Splitter	405, 488, 561, 642
Filter Excitation Wavelength	200 - 405, 200 - 488, 200 - 561, 200 - 642
Filter Emission Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Contrast Method	Fluorescence
Illumination Wavelength	561 405
Channel Name	T1-Axiocam 820 #2-SR T3-Axiocam 820 #2-SR
Excitation Wavelength	561 405
Emission Wavelength	597 445
Effective NA	1.4
Detection Wavelength	419-470, 503-546, 576-617, 657-800
Imaging Device	Axiocam 820 #2
Camera Adapter	1X
Exposure Time	120ms 50ms
Depth of Focus	0.92µm 0.69µm
Binning Mode	2,2
Laser Wavelength	561nm @2.00% 405nm @7.0%
Frame Time	1.33s 676.00ms
Scan Direction	Unidirectional
Result Sampling	4
Process Sampling	N/a
Algorithm	SIM

FIELD	SOURCE-DOCUMENTED VALUE
SIM Settings	Standard
Auto Sharpness	On
Sharpness	10.3593 (594), 8.7073 (DAPI)
Fitting	Fast fit
Histogram	Scale to Min/Max
Baseline	Cut
3D Maximum Projection	Precise
Appearance	Threshold 4% (594), Threshold 4% (405) Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 127%, Enabled Tone Mapping
Projection	View Angle 51°, Scale Z 20%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

HeLa Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 91 Slices (4.50µm)

Channels = 2

Scaling (Per Pixel) = 16.89nm X 16.89nm X 50.00nm

Image Size (Pixels) = 5056 X 5056

Image Size (Scaled) = 85.40µm X 85.40µm

Bit Depth = 16 Bit

Image Center Position = X: 58.12mm, Y: 32.33mm

Stage Position = X: 58.12mm, Y: 32.33mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

594 -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 561
 Channel Name = T1-Axiocam 820 #2-SR
 Excitation Wavelength = 561
 Emission Wavelength = 597
 Effective NA = 1.4
 Detection Wavelength = 419-470, 503-546, 576-617, 657-800
 Imaging Device = Axiocam 820 #2
 Camera Adapter = 1X
 Exposure Time = 120ms
 Depth of Focus = 0.92µm
 Binning Mode = 2,2
 Laser Wavelength = 561nm @2.00%
 Frame Time = 1.33s
 Scan Direction = Unidirectional
 Grating 36.5µm grating

DAPI -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 405
 Channel Name = T3-Axiocam 820 #2-SR
 Excitation Wavelength = 405
 Emission Wavelength = 445
 Effective NA = 1.4
 Detection Wavelength = 419-470, 503-546, 576-617, 657-800
 Imaging Device = Axiocam 820 #2
 Camera Adapter = 1X
 Exposure Time = 50ms
 Depth of Focus = 0.69µm
 Binning Mode = 2,2
 Laser Wavelength = 405nm @7.0%
 Frame Time = 676.00ms
 Scan Direction = Unidirectional
 Grating 36.5µm grating

Post Processing - SIM Processing

Processing Dimension - 3D

Result Sampling = 4

Process Sampling = N/a

Channels = 2

Algorithm = SIM

SIM Settings = Standard

Auto Sharpness = On

Sharpness = 10.3593 (594), 8.7073 (DAPI)

Fitting = Fast fit

Histogram = Scale to Min/Max

Baseline = Cut

3D Image Processing

3D Maximum Projection = Precise

Appearance = Threshold 4% (594), Threshold 4% (405) Ramp 0% all channels, Maximum 100% all channels

Background = Black

Light = Brightness 127%, Enabled Tone Mapping

Projection = View Angle 51°, Scale Z 20%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera

Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections


None

Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000051
Stage	SIM
Instrument	ZEISS Elyra
Microscope configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Structured source metadata values	50
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	48455989317023EB7BA72AB4E96DCC40B9D2ECEAC23DFA8FEDFDB6164EF10460
Method PDF SHA-256	DFB3C6354ACAF6973D0F8FEBE09CFC57DDBA89D5C19BC433875A388387685E4D

ORIGINAL IMAGE EVIDENCE / SIM²

EVIDENCE TIER	SIM ²
INSTRUMENT	ZEISS Elyra 7
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	57

STRUCTURED ACQUISITION METADATA / SIM²

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 7, Andor1, AxioObserver, using Plan-Apochromat 63X/1.46 oil Korr M27 was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.46 oil Korr M27
Filters	-2147483648
Beam Splitter 594	Main Beam Splitter Non-Descanned: LBF 405/488/561/642
Dual Camera Beam Splitter	SBS LP 640
Beam Splitter 405	Main Beam Splitter Non-Descanned: LBF 405/488/561/642
Calibration Marker Positions	Marker 1: 0.00µm, 0.00µm, Marker 2: 0.00µm, 0.00µm, Marker 3 0.00µm, 0.00µm
Z-Stack	28 Slices (1.62µm)
Channels	2
Scaling (Per Pixel)	0.01565 µm X 0.01565 µm X 0.06000 µm
Image Size (Pixels)	3979 X 4207
Image Size (Scaled)	62.68µm X 65.85µm
Bit Depth	16 Bit
Image Center Position	X: 8.80mm, Y: 6.45mm
Stage Position	X: 8.80mm, Y: 6.45mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Contrast Method	Fluorescence
Laser Wavelength	561nm @3.5% 405nm @4.0%
Scan Mode	Other
Scan Zoom	1.0
Rotation	0°
Frame Time	1.0s
Scan Direction	Unidirectional
Line Step	N/A
Averaging	8
Excitation Wavelength	561 405
Emission Wavelength	515
Effective NA	1.46
Imaging Device	Edge 4.2 SIM
Exposure Time	100ms 50ms
Depth of Focus	0.73µm
Binning Mode	1x1
Detector Type	Camera

FIELD	SOURCE-DOCUMENTED VALUE
Depth Offset	0
Detector Digital Gain	0.00
Channel	561.000000 405.000000
Grating Period (µm)	27.500000 36.500000
Processing	3D
Input SNR	Medium
Iterations	16
Regularization Weight	0.065000
Processing Sampling	4
Output Sampling	4
Filter	Median
Detrend	No
Sectioning	100.000000
Baseline	Yes
Scale to Raw Image	No
Experimental PSF	No
3D Maximum Projection	Precise
Appearance	Threshold 1% (594), Threshold 1% (405) Ramp 0% (594), Ramp 0% (405), Maximum 100% all channels
Background	Black
Light	Brightness 125%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Mitochondria Antibody (MA5-12017), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Elyra 7, Andor1, AxioObserver, using Plan-Apochromat 63X/1.46 oil Korr M27 was used with the following parameters for each dye:

Filters = -2147483648

Beam Splitter 594 = Main Beam Splitter Non-Descanned: LBF 405/488/561/642

Dual Camera Beam Splitter = SBS LP 640

Beam Splitter 405 = Main Beam Splitter Non-Descanned: LBF 405/488/561/642

Dual Camera Beam Splitter = SBS LP 640

Calibration Marker Positions: Marker 1: 0.00µm, 0.00µm, Marker 2: 0.00µm, 0.00µm, Marker 3 0.00µm, 0.00µm

Z Stack - (3D)

Z-Stack = 28 Slices (1.62µm)

Channels = 2

Scaling (Per Pixel) = 15.65nm X 15.65nm X 60.00nm

Image Size (Pixels) = 3979 X 4207

Image Size (Scaled) = 62.68µm X 65.85µm

Bit Depth = 16 Bit

Image Center Position = X: 8.80mm, Y: 6.45mm

Stage Position = X: 8.80mm, Y: 6.45mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

594 -

Contrast Method = Fluorescence
 Laser Wavelength = 561nm @3.5%
 Scan Mode = Other
 Scan Zoom = 1.0
 Rotation = 0°
 Frame Time = 1.0s
 Scan Direction = Unidirectional
 Line Step = N/A
 Averaging = 8
 Excitation Wavelength = 561
 Emission Wavelength = 515
 Effective NA = 1.46
 Imaging Device = Edge 4.2 SIM
 Exposure Time = 100ms
 Depth of Focus = 0.73µm
 Binning Mode = 1x1
 Detector Type = Camera
 Depth Offset = 0
 Detector Digital Gain = 0.00

Post Processing - SIM2 Processing

Channel = 561.000000
Grating Period (µm): 27.500000

SIM2

Processing: 3D
 Input SNR: Medium
 Iterations: 16
 Regularization Weight: 0.065000
 Processing Sampling: 4
 Output Sampling: 4
 Filter: Median
 Detrend: No
 Sectioning: 100.000000
 Baseline: Yes
 Scale to Raw Image: No
 Experimental PSF: No

DAPI -

Contrast Method = Fluorescence
 Laser Wavelength = 405nm @4.0%
 Scan Mode = Other
 Scan Zoom = 1.0
 Rotation = 0°
 Frame Time = 1.0s
 Scan Direction = Unidirectional
 Line Step = N/A
 Averaging = 8
 Excitation Wavelength = 405
 Emission Wavelength = 515
 Effective NA = 1.46
 Imaging Device = Edge 4.2 SIM
 Exposure Time = 50ms
 Depth of Focus = 0.73µm
 Binning Mode = 1x1
 Detector Type = Camera
 Depth Offset = 0
 Detector Digital Gain = 0.00

Post Processing - SIM2 Processing

Channel = 405.000000
Grating Period (µm): 36.500000

SIM2

Processing: 3D
 Input SNR: Medium
 Iterations: 16
 Regularization Weight: 0.065000
 Processing Sampling: 4
 Output Sampling: 4
 Filter: Median
 Detrend: No
 Sectioning: 100.000000
 Baseline: Yes
 Scale to Raw Image: No
 Experimental PSF: No

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 1% (594), Threshold 1% (405) Ramp 0% (594), Ramp 0% (405), Maximum 100% all channels
 Background = Black
 Light = Brightness 125%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None


Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000051
Stage	SIM²
Instrument	ZEISS Elyra 7
Microscope configuration	Zeiss Elyra 7, Andor1, AxioObserver, using Plan-Apochromat 63X/1.46 oil Korr M27 was used with the following parameters for each dye:
Structured source metadata values	57
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	7D2FE9BCEFC6557DAAC99B956EDE3DEC75CEA7F198A9B0D9FA77FF9873CA3ABC
Method PDF SHA-256	5DDF23A5D23FFE0E1C919A5FDEAF297993E852A23819F3F188E4DAB8E3AFD2CD

ORIGINAL IMAGE EVIDENCE / CONTROL



EVIDENCE TIER	Control
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 594
METADATA VALUES	32

STRUCTURED ACQUISITION METADATA / CONTROL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Channels	2
Scaling (Per Pixel)	0.143µm X 0.143µm
Image Size (Pixels)	1183 X 872
Image Size (Scaled)	169.00µm X 124.57µm
Bit Depth	14 Bit
Image Center Position	X: 50.68mm, Y: 50.48mm
Stage Position	X: 50.68mm, Y: 50.48mm
ROI Center Offset	X: 1.14µm, Y: 0.00µm
Reflector	45 Texas Red 49 DAPI
Beam Splitter	585 395
Filter Excitation Wavelength	540 - 580 335 - 383
Filter Emission Wavelength	593 - 668 420 - 470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @25% X-Cite Xylis Lamp Intensity @10%
Exposure Time	200ms 30ms
Excitation Wavelength	590 353
Emission Wavelength	618 465
Effective NA	1.25
Imaging Device	Axiocam 807 mono
Camera Adapter	1X
Depth of Focus	1.20µm 0.90µm
Binning Mode	2,2
Apotome SIM Image Processing	Display Mode "Optical Sectioning" - enabled correction selected.

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Goat Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000051

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. In the absence of a primary antibody, the Identifyn® 594 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.143µm X 0.143µm

Image Size (Pixels) = 1183 X 872

Image Size (Scaled) = 169.00µm X 124.57µm

Bit Depth = 14 Bit

Image Center Position = X: 50.68mm, Y: 50.48mm

Stage Position = X: 50.68mm, Y: 50.48mm

ROI Center Offset = X: 1.14µm, Y: 0.00µm

594 -

Reflector = 45 Texas Red

Beam Splitter = 585

Filter Excitation Wavelength = 540 - 580

Filter Emission Wavelength = 593 - 668

Contrast Method = Fluorescence

Light source = X-Cite Xylis Lamp Intensity @25%

Exposure Time = 200ms

Excitation Wavelength = 590

Emission Wavelength = 618

Effective NA = 1.25

Imaging Device = Axiocam 807 mono

Camera Adapter = 1X

Depth of Focus = 1.20µm

Binning Mode = 2,2

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335 - 383
 Filter Emission Wavelength = 420 - 470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @10%
 Exposure Time = 30ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 0.90µm
 Binning Mode = 2,2

Post Processing - SIM Processing

Apotome SIM Image Processing = Display Mode "Optical Sectioning" - enabled correction selected.

Normalization, Phase Corrector, Fourier Filter, and Deconvolution were NOT used.

Image Corrections

None

Control Summary -

In the absence of a primary antibody, Identifyn Standards for Optical Microscopy Methods™ were applied using standard staining methodology and image acquisition settings using the Zeiss Apotome SIM. With no image correction, the image demonstrates that the secondary antibody, minus a primary target, had no off-target binding and no significant presence of free dye within the sample.

The control data suggests the secondary antibody is highly specific and free of unconjugated (free) dye post-conjugation.

Image by J.K. Bennett

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000051
Stage	Control
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	32

SOURCE PROVENANCE

Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	B08FE13416BF8AD6CD7553B19E104F22E840790E84A412A2188A4361809EF85D
Method PDF SHA-256	D937B9D5BBACBC92E194683F50794258AD77C57F5C1247D07D0598DA3CB377A0