

IDENTIFYN RESOLUTION STEPDOWN

SECONDARY ANTIBODY MICROSCOPY EVIDENCE ASSESSMENT

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human, IgG (H + L)

WESTERN BLOT -> WIDEFIELD -> WIDEFIELD SIM / APOTOME -> CONFOCAL -> AIRYSCAN
SIM -> SIM² -> SINGLE-MOLECULE STORM

8


ORDERED EVIDENCE TIERS

1

SUPPORTING CONTROL

PENDING

SCIENTIST REVIEW



REPRESENTATIVE HIGHEST-AVAILABLE VIEW / SINGLE-MOLECULE STORM

SA-000014 | PRODUCT-LEVEL EVIDENCE DOSSIER

SOURCE-GROUNDED MICROSCOPY + BIOLOGICAL REVIEW / IDENTIFYN EVIDENCE PRESERVED

CRO CAPABILITY DEMONSTRATED

Identifyn performed the experimental and imaging work documented below. BENNETT converts that operating history into a governed, source-bounded scientific record.

- Source-documented sample preparation and antibody/reagent workflow
- Western blot imaging
- Widefield fluorescence microscopy
- Widefield structured illumination / Apotome optical sectioning
- Confocal microscopy
- Airyscan acquisition and reconstruction workflow
- Structured illumination microscopy
- SIM² reconstruction workflow
- Single-molecule STORM presentation workflow
- Preserved control experiment
- Treatment-response experimental workflow
- Z-stack / 3D acquisition
- Multi-channel imaging

SCIENTIFIC QUESTION

What source-bounded experimental and microscopy evidence is preserved for Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human, IgG (H + L), and what is the strongest interpretation allowed by the available modalities and controls?

EVIDENCE AVAILABLE

Western blot -> Widefield -> Widefield SIM — Apotome -> Confocal -> Airyscan -> SIM -> SIM² -> Single-molecule STORM -> Control

BENNETT ASSESSMENT

The preserved SA-000014 record contains the full ordered evidence sequence from Western blot through single-molecule STORM. Each modality retains its own acquisition and processing boundary.

BENNETT / MICROSCOPY AND BIOLOGICAL EVIDENCE REVIEW

SOURCE-GROUNDED RETROSPECTIVE REVIEW / PENDING SCIENTIST REVIEW

SOURCE STATE	SOURCE HASH VERIFIED / EVENT RECORDED
BENNETT EVIDENCE REVIEW	COMPLETED / SOURCE-BOUNDED
NATIVE IMAGE REPROCESSING	NOT PERFORMED
LITERATURE SOURCES	INDEPENDENTLY VERIFIED
SCIENTIST REVIEW	PENDING
PUBLICATION	NOT AUTHORIZED

OVERALL CALL / FULL STEPDOWN / SOURCE-BOUNDED REVIEW

The preserved SA-000014 record contains the full ordered evidence sequence from Western blot through single-molecule STORM. Each modality retains its own acquisition and processing boundary.

SETTINGS ASSESSMENT

The record preserves 8 ordered evidence tier(s). Structured source metadata values are reported by stage and remain tied to the original method objects.

BIOLOGICAL SUPPORT

The record supports retrospective, source-bounded microscopy review. Target identity, specificity, treatment effect, binding, interaction and mechanism are not inferred from presentation images alone.

CONTROL ASSESSMENT

A true preserved control record is present and appears last.

CLAIM CEILING

This draft supports historical capability documentation, source registration, modality ordering, settings review and bounded interpretation. It does not independently validate antibody specificity, quantify a population response, establish binding, interaction, causation or mechanism, or authorize publication.

PROCESSING DISCLOSURE

Retrospective BENNETT architectural evaluation of preserved Identifyn source evidence. BENNETT did not perform native-file reprocessing, new deterministic measurement or the historical image post-processing represented in this record.

REAGENTS AND ACQUIRED CHANNELS

Preparation reagents are reported separately from channels actually documented in the acquisition settings. Fluorophores mentioned in staining are not automatically treated as acquired channels.

EVIDENCE OBJECT	REAGENTS / FLUOROPHORES PRESENT	CHANNELS ACTUALLY ACQUIRED
Western blot	488	Not deterministically stated in the acquisition block
Widefield	405/DAPI; 488; 594	2; 38 HE eGFP; 45 Texas Red; 450 - 490; 540 - 580; 500 - 550; 593 - 668; 493
Widefield SIM — Apotome	405/DAPI; 488; 594	3; 38 HE eGFP; 45 Texas Red; 49 DAPI; 450 - 545490; 540 - 580; 335-383; 500 - 550
Confocal	405/DAPI; 488; 594	2; None; 488nm @0.2%; 561nm @0.2%; 493; 590; 517; 618
Airyscan	405/DAPI; 488	2; None; 488nm @0.2%; 405nm @1.50%; 493; 353; 517; 465
SIM	405/DAPI; 488; 594	2; 1; SR HE LBF 405/488/561/642; 200 - 405, 200 - 488, 200 - 561, 200 - 642; 419 - 470, 503 - 546, 576 - 617, 657 - 800; T2-Axiocam 820 -SR; T1-Axiocam 820 #2-SR; 488
SIM ²	405/DAPI; 488; 594	2; 1; SR HE LBF 405/488/561/642; 200 - 405, 200 - 488, 200 - 561, 200 - 642; 419 - 470, 503 - 546, 576 - 617, 657 - 800; T2-Axiocam 820 -SR; T1-Axiocam 820 #2-SR; 488
Single-molecule STORM	488	Both Channels
Control	405/DAPI; 488	2; 38 HE eGFP; 49 DAPI; 450 - 490; 335-383; 500 -550; 420-470; 493

MODALITY ASSESSMENT / PART 1 OF 9

Western blot / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: Azure Sapphire FL. Objective: not explicitly stated. Platform: Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s). Structured source metadata values: 11.

VISIBLE OBSERVATION

A preserved presentation image is available for Western blot. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Western blot, documents platform context as Azure Sapphire FL, and records detected dye/channel descriptors as 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Band position and displayed intensity do not independently establish analytical specificity, target identity, linearity, or treatment effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Western blot evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 2 OF 9

Widefield / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Imager.M1. Objective: Plan-Apochromat 63x/1.40 Oil M27. Platform: Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Channels: 2; Scaling (Per Pixel): 0.110µm X 0.110µm; Image size (Pixels): 1216 X 1028; Image Size (Pixels): 1216 X 1028; Bit Depth: 14 Bit. Detector/camera: Imaging Device: Axiocam MR R3. Timing: Exposure Time: 180ms; 80ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @30%. Optical/scan: Effective NA: 1.4. Structured source metadata values: 30.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield, documents platform context as ZEISS Axio Imager.M1, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

A presentation image supports gross localization and source-bounded co-occurrence, not direct binding, molecular colocalization, or native quantitative measurement.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 3 OF 9

Widefield SIM — Apotome / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Z-Stack: 11 Slices (2.0µm); Channels: 3; Scaling (Per Pixel): 0.143µm X 0.143µm X 0.200µm; Image size (Pixels): 1216 X 991; Image Size (Pixels): 1216 X 991; Bit Depth: 14 Bit. Detector/camera: Imaging Device: AxioCam 807. Timing: Exposure Time: 150ms; 100ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @30%; X-Cite Xylis Lamp Intensity @6.0%. Optical/scan: Effective NA: 1.25. Processing: Projection: View Angle 35°, Scale Z 182%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 45.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield SIM — Apotome. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield SIM — Apotome, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Improved optical sectioning does not by itself establish reconstructed super-resolution or molecular interaction.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield SIM — Apotome evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 4 OF 9

Confocal / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 70 Slices (4.41µm); Channels: 2; Scaling (Per Pixel): 0.071µm X 0.071µm X 0.060µm; Image size (Pixels): 1214 X 1456; Image Size (Pixels): 1214 X 1456; Bit Depth: 16 Bit. Detector/camera: Imaging Device: GaAsP-PMT 1; GaAsP-PMT 2; Detector Type: GaAsP-PMT. Timing: Pixel Time: 0.51µs; Frame Time: 10.15s. Illumination: Laser Wavelength: 488nm @0.2%; 561nm @0.2%. Optical/scan: Pinhole: 1.00AU / 45µm; 1.00AU / 54µm; Scan Zoom: 0.70; LSM Scan Speed: 7; Scan Direction: Bidirectional; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 44.

VISIBLE OBSERVATION

A preserved presentation image is available for Confocal. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Confocal, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Optical sectioning and source-bounded 3D presentation do not establish binding, mechanism, or population-level effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Confocal evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 5 OF 9

Airyscan / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 49 Slices (2.4µm); Channels: 2; Scaling (Per Pixel): 35.29nm X 35.29nm X 50nm; Image size (Pixels): 1262 X 670; Image Size (Pixels): 1262 X 670; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Airyscan; Detector Type: GaAsp-PMT. Timing: Pixel Time: 0.59µs; Frame Time: 4.44s. Illumination: Laser Wavelength: 488nm @0.2%; 405nm @1.50%. Optical/scan: Pinhole: 5.00 AU / 219µm; 5.00 AU / 181µm; Scan Zoom: 1.6; LSM Scan Speed: 7; Scan Direction: Bidirectional; Averaging: 2; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 90%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black; AiryScan Mode: Super Resolution 7.9 (3D, Auto); Super Resolution 7.6 (3D, Auto). Structured source metadata values: 46.

VISIBLE OBSERVATION

A preserved presentation image is available for Airyscan. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Airyscan, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Displayed feature separation remains reconstruction dependent; no unsupported numerical resolution or molecular architecture claim is permitted.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Airyscan evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 6 OF 9

SIM / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:. Acquisition: Z-Stack: 35 Slices (1.70µm); Channels: 2; 1; Scaling (Per Pixel): 13.20nm X 13.20nm X 50.00nm; Image size (Pixels): 5451 X 5106; Image Size (Pixels): 5451 X 5106; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Axiocam 820; Axiocam 820#2. Timing: Exposure Time: 100ms; 120ms; Frame Time: 1.33s. Illumination: Laser Wavelength: 488nm @1.50%; 561nm @3.0%; Illumination Wavelength: 488; 561. Optical/scan: Scan Direction: Unidirectional; Effective NA: 1.4. Processing: Algorithm: SIM; Process Sampling: 2; Result Sampling: 4; Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 55.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM, documents platform context as ZEISS Elyra, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Reconstruction-dependent structural separation requires native reconstruction QC before numerical resolution or molecular dimensions can be claimed.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 7 OF 9

SIM² / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:. Acquisition: Z-Stack: 39 Slices (1.90µm); Channels: 2; 1; Scaling (Per Pixel): 13.20nm X 13.20nm X 50.00nm; Image size (Pixels): 2324 X 2471; Image Size (Pixels): 2324 X 2471; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Axiocam 820; Axiocam 820#2. Timing: Exposure Time: 100ms; 120ms; Frame Time: 1.33s. Illumination: Laser Wavelength: 488nm @1.50%; 561nm @3.0%; Illumination Wavelength: 488; 561. Optical/scan: Scan Direction: Unidirectional; Effective NA: 1.4. Processing: Algorithm: SIM2; Process Sampling: 4; Result Sampling: 4; Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 55.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM². BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM², documents platform context as ZEISS Elyra, and records detected dye/channel descriptors as 405/DAPI; 488; 594.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Further reconstruction-dependent segmentation does not establish molecular dimensions, structure identity, or mechanism without independent support.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM² evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 8 OF 9

Single-molecule STORM / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: Bruker Vutara VXL. Objective: Olympus 60X/1.30 silicon oil objective. Platform: Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:.

Detector/camera: Camera: Both Cameras. Timing: Exposure Time: 20ms Selected; Super Resolution Camera Exposure Time: 20ms; Widefield Camera Exposure Time: 100ms. Illumination: Photoactivation/Imaging Laser: 0% Selected.

Structured source metadata values: 71.

VISIBLE OBSERVATION

A preserved presentation image is available for Single-molecule STORM. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Single-molecule STORM, documents platform context as Bruker Vutara VXL, and records detected dye/channel descriptors as 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Without localization tables, uncertainty, blinking, drift, and independent cluster evidence, the presentation cannot establish molecule count, cluster truth, exact precision, or molecular architecture.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Single-molecule STORM evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 9 OF 9

Control / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Channels: 2; Scaling (Per Pixel): 0.143µm X 0.143µm; Image size (Pixels): 1422 X 960; Image Size (Pixels): 1422 X 960; Bit Depth: 14 Bit. Detector/camera: Imaging Device: AxioCam 807 mono. Timing: Exposure Time: 200ms; 30ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @25%; X-Cite Xylis Lamp Intensity @10%. Optical/scan: Effective NA: 1.25. Structured source metadata values: 30.

VISIBLE OBSERVATION

A preserved presentation image is available for Control. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Control, documents platform context as ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 488.

LITERATURE CORRELATION

Consistent with peer-reviewed antibody-control guidance that a no-primary or related control bounds background but is not a complete specificity system.

INTERPRETIVE LIMIT

A control bounds the recorded comparison only; it does not establish universal specificity, zero background, zero cross-reactivity, or absence of free dye.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Control evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

CROSS-MODALITY SYNTHESIS

The preserved SA-000014 record contains the full ordered evidence sequence from Western blot through single-molecule STORM. Each modality retains its own acquisition and processing boundary.

Different acquisitions are not treated as a quantitative intensity ladder. Any continuity statement remains qualitative and source bounded.

BENNETT LITERATURE CORRELATION — SOURCES INDEPENDENTLY VERIFIED

BENNETT uses independently verified primary scientific sources to test whether the interpretation is scientifically consistent and properly bounded. Literature correlation does not retroactively validate the Identifyn dataset or replace scientist review.

- Antibody and microscopy evidence require orthogonal controls and modality-appropriate interpretation; presentation images alone do not establish analytical specificity.
- Advanced-resolution presentation depends on acquisition, reconstruction or localization quality control before numerical resolution, molecular dimensions or cluster identity can be claimed.

REFERENCES

1. Uhlen M, et al. A proposal for validation of antibodies. Nat Methods. 2016;13:823-827. doi:10.1038/nmeth.3995.
2. Gustafsson MGL. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. J Microsc. 2000;198:82-87. doi:10.1046/j.1365-2818.2000.00710.x.
3. Sage D, et al. Quantitative evaluation of software packages for single-molecule localization microscopy. Nat Methods. 2015;12:717-724. doi:10.1038/nmeth.3442.

RELEASE BOUNDARY

This assessment remains a draft pending scientist review. No publication is authorized. Any future native-file analysis must enter BENNETT as a new evidence snapshot with routed engine authority, native-data QA, methods, limitations, and an independently reviewed claim ceiling.

SOURCE RECONCILIATION / QA / PROVENANCE

SOURCE STATUS - PRESERVED / AUTO-RECONCILED

Historical Identifyn source material was subjected to BENNETT retrospective QA. Original source state and correction history remain preserved in provenance.

HIGH CONFIDENCE - AUTO-RECONCILED

Airyscan / Scaling (Per Pixel): 35.29nm X 35.29nm X 50nm -> 0.03529 μ m X 0.03529 μ m X 0.05000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=1262 X 670; Image Size (Scaled)=44.54 μ m X 23.65 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.02%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM / Scaling (Per Pixel): 13.20nm X 13.20nm X 50.00nm -> 0.01320 μ m X 0.01320 μ m X 0.05000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=5451 X 5106; Image Size (Scaled)=71.93 μ m X 67.38 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.03%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM² / Scaling (Per Pixel): 13.20nm X 13.20nm X 50.00nm -> 0.01320 μ m X 0.01320 μ m X 0.05000 μ m. Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=2324 X 2471; Image Size (Scaled)=30.67 μ m X 32.61 μ m; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.02%.

UNRESOLVED - PRESERVED AS CONFLICT

Cross-tier / treatment duration: {"Widefield": ["24 hours"], "Widefield SIM — Apotome": ["24 hours"], "Confocal": ["24 hours"], "Airyscan": ["24 hours"], "SIM": ["24 hours"], "SIM²": ["24 hours"], "Single-molecule STORM": ["6 hours"]} -> Preserved as modality-specific historical duration values. Genuinely different treatment durations are present across evidence stages, and no source or scientist correction resolves them. Supporting evidence: Only duration values from explicit etoposide-treatment sentences were classified; concentration, wavelength, resolution and pixel-size values were excluded.

BENNETT EVIDENCE REVIEW COMPLETE

INTERPRETATION TO EVIDENCE

BENNETT has completed the bounded retrospective assessment.
The following pages switch ownership to the preserved Identifyn record.



IDENTIFYN PRESERVED SOURCE EVIDENCE

ORIGINAL RECORD BEGINS

Images, source methods, acquisition settings and provenance follow in the locked
Stepdown order. Scientific wording is preserved; missing modalities are not invented.
A preserved control follows all available Stepdown tiers as supporting evidence.

PRODUCT-LEVEL STEPDOWN MAP


This dossier preserves the complete available evidence progression for SA-000014. Each tier remains tied to its original image, source method, instrument context and cryptographic identity. A preserved control is retained as supporting evidence and appears after every resolution tier.

CANONICAL STEPDOWN	EVIDENCE STAGE	INSTRUMENT	STATUS
1	Western blot	Azure Sapphire FL	PRESERVED
2	Widefield	ZEISS Axio Imager.M1	PRESERVED
3	Widefield SIM - Apotome	ZEISS Axio Observer.Z1/7 Apotome	PRESERVED
4	Confocal	ZEISS LSM 900	PRESERVED
5	Airyscan	ZEISS LSM 900	PRESERVED
6	SIM	ZEISS Elyra	PRESERVED
7	SIM ²	ZEISS Elyra	PRESERVED
8	Single-molecule STORM	Bruker Vutara VXL	PRESERVED

INTERPRETIVE BOUNDARY

Ordering evidence by resolution tier does not make the modalities interchangeable. Each stage retains its acquisition environment and scientific limitations. This pilot organizes the record; it does not synthesize new biological claims.

ORIGINAL IMAGE EVIDENCE / WESTERN BLOT



EVIDENCE TIER	Western blot
INSTRUMENT	Azure Sapphire FL
CELL MODEL	HeLa
DETECTED DYES	488
METADATA VALUES	11

STRUCTURED ACQUISITION METADATA / WESTERN BLOT

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Scanner	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Detection mode	Fluorescence
Laser	488nm
Emission Filter	513±17nm
Detector	PMT
Intensity	L4
Pixel size	100 µm
Scan Speed	High
Quality	High
Focus	Membrane (0 µm in Z-axis)
Image	Identifyn® LLC

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Etoposide-treated (2 µM) HeLa cells were homogenized using an ultrasonic probe and centrifuged to pellet the cell debris. The lysate supernatant was mixed with NuPAGE™ LDS Sample Buffer (4X), heated at 80 °C for 10 min, and subjected to electrophoresis using NuPAGE™ 4 to 12%, Bis-Tris, 1.0-1.5 mm, Mini Protein Gels in a Mini Gel Tank. The Thermo Fisher's PageRuler™ Prestained Protein Ladder, 10 to 180 kDa, was used as a molecular weight marker.

Post electrophoresis, the proteins were transferred to a polyvinylidene difluoride (PVDF) membrane using the iBlot™ 3 Starter Kit, PVDF. The transfer membrane was blocked with Blocker™ Casein for 1 hour and then washed five times with PBS. The blocked membrane was incubated with Identifyn® Rabbit Anti-Human γH2AX chimeric antibody (Human Fc) for 2 hours, followed by five washes in 1X PBS. Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat anti-Human IgG was incubated for 1 hour, followed by washing 5X with PBS. The membrane was dried for 10 minutes at room temperature before imaging.

Scanner = Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

Scan Settings

Detection mode = Fluorescence
Laser = 488nm
Emission Filter = 513±17nm
Detector = PMT
Intensity = L4
Pixel size = 100 µm
Scan Speed = High
Quality = High
Focus = Membrane (0 µm in Z-axis)

Post Processing

None

Image: Identifyn® LLC

Copyright 2026 GMD12 LLC.
Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000014
---------	-----------

SOURCE PROVENANCE

Stage	Western blot
Instrument	Azure Sapphire FL
Microscope configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Structured source metadata values	11
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	AC6B86CD3A3DABC66F4419BB497DCFDA546AE21E975F1833C8F5BCA54741AEE
Method PDF SHA-256	C255F42F61CF4B53626A58C6810EDFC629D5F11F8512020955BF4BA421217A67

ORIGINAL IMAGE EVIDENCE / WIDEFIELD



EVIDENCE TIER	Widefield
INSTRUMENT	ZEISS Axio Imager.M1
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	30

STRUCTURED ACQUISITION METADATA / WIDEFIELD

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	Plan-Apochromat 63x/1.40 Oil M27
Channels	2
Scaling (Per Pixel)	0.110µm X 0.110µm
Image size (Pixels)	1216 X 1028
Image Size (Scaled)	133.18µm X 112.59µm
Bit Depth	14 Bit
Image Center Position	X: 0.00µm, Y: 0.00µm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	38 HE eGFP 45 Texas Red
Beam Splitter	495 585
Filter Excitation Wavelength	450 - 490 540 - 580
Filter Emission Wavelength	500 - 550 593 - 668
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @30%
Exposure Time	180ms 80ms
Excitation Wavelength	493 590
Emission Wavelength	517 618
Effective NA	1.4
Imaging Device	Axiocam MR R3
Camera Adapter	1X
Depth of Focus	0.80µm 0.96µm
Binning Mode	2,2 12,2

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000048

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Mitochondria Antibody (MA5-12017), followed, and was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 594 secondary antibody was incubated for 1 hour at room temperature at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.110 μ m X 0.110 μ m

Image size (Pixels) = 1216 X 1028

Image Size (Scaled) = 133.18 μ m X 112.59 μ m

Bit Depth = 14 Bit

Image Center Position = X: 0.00 μ m, Y: 0.00 μ m

ROI Center Offset = X: 0.00 μ m, Y: 0.00 μ m

488 -

Reflector = 38 HE eGFP
 Beam Splitter = 495
 Filter Excitation Wavelength = 450 - 490
 Filter Emission Wavelength = 500 - 550
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @30%
 Exposure Time = 180ms
 Excitation Wavelength = 493
 Emission Wavelength = 517
 Effective NA = 1.4
 Imaging Device = Axiocam MR R3
 Camera Adapter = 1X
 Depth of Focus = 0.80µm
 Binning Mode = 2,2

594 -

Reflector = 45 Texas Red
 Beam Splitter = 585
 Filter Excitation Wavelength = 540 - 580
 Filter Emission Wavelength = 593 - 668
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @30%
 Exposure Time = 80ms
 Excitation Wavelength = 590
 Emission Wavelength = 618
 Effective NA = 1.4
 Imaging Device = Axiocam MR R3
 Camera Adapter = 1X
 Depth of Focus = 0.96µm
 Binning Mode = 12,2

Post Processing

None

Image Correction

None

Image by K. Shivanna
 Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000014
Stage	Widefield

SOURCE PROVENANCE

Instrument	ZEISS Axio Imager.M1
Microscope configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam MR R3, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	30
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	950C67331F03E13A22E3C57833C8818A3BC32EF4EA79D65649A94DDCEACAFBC6
Method PDF SHA-256	CE7236F06AA7C8E23ED5E6C04FEE7EA124373EC32307038E40AFE58894AC40FC

ORIGINAL IMAGE EVIDENCE / WIDEFIELD SIM - APOTOME



EVIDENCE TIER	Widefield SIM - Apotome
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	45

STRUCTURED ACQUISITION METADATA / WIDEFIELD SIM - APOTOME

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Z-Stack	11 Slices (2.0µm)
Channels	3
Scaling (Per Pixel)	0.143µm X 0.143µm X 0.200µm
Image Size (Pixels)	1216 X 991
Image Size (Scaled)	173.71µm X 141.57µm
Bit Depth	14 Bit
Image Center Position	X: 88.46mm, Y: 47.28mm
Stage Position	X: 88.46mm, Y: 47.28mm
ROI Center Offset	X: 1.14, Y: 0.00µm
Reflector	38 HE eGFP 45 Texas Red 49 DAPI
Beam Splitter	495 585 395
Filter Excitation Wavelength	450 - 545490 540 - 580 335-383
Filter Emission Wavelength	500 - 550 593 - 668 420-470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @30% X-Cite Xylis Lamp Intensity @6.0%
Exposure Time	150ms 100ms 18ms
Excitation Wavelength	493 590 353
Emission Wavelength	517 618 465
Effective NA	1.25
Imaging Device	Axiocam 807
Camera Adapter	1X
Depth of Focus	1.00µm 0.90µm
Binning Mode	2,2
Apotome SIM Image Processing	Display Mode "Optical Sectioning" - enabled correction selected.
3D Volume Projection	Precise
Appearance	Threshold 13% (488), Threshold 3% (594), Threshold13% (DAPI), Ramp 50% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 120%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, did NOT Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 182%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000048

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Mitochondria Antibody (MA5-12017), was then applied, and the sample was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 594 secondary antibody was incubated for 1 hour at room temperature at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI, cat#P36962.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 11 Slices (2.0µm)

Channels = 3
Scaling (Per Pixel) = 0.143µm X 0.143µm X 0.200µm
Image Size (Pixels) = 1216 X 991
Image Size (Scaled) = 173.71µm X 141.57µm
Bit Depth = 14 Bit
Image Center Position = X: 88.46mm, Y: 47.28mm
Stage Position = X: 88.46mm, Y: 47.28mm
ROI Center Offset = X: 1.14, Y: 0.00µm

488 -

Reflector = 38 HE eGFP
Beam Splitter = 495
Filter Excitation Wavelength = 450 - 545490
Filter Emission Wavelength = 500 - 550
Contrast Method = Fluorescence
Light source = X-Cite Xylis Lamp Intensity @30%
Exposure Time = 150ms
Excitation Wavelength = 493
Emission Wavelength = 517
Effective NA = 1.25
Imaging Device = AxioCam 807
Camera Adapter = 1X
Depth of Focus = 1.00µm
Binning Mode = 2,2

594 -

Reflector = 45 Texas Red
Beam Splitter = 585
Filter Excitation Wavelength = 540 - 580
Filter Emission Wavelength = 593 - 668
Contrast Method = Fluorescence
Light source = X-Cite Xylis Lamp Intensity @30%
Exposure Time = 100ms
Excitation Wavelength = 590
Emission Wavelength = 618
Effective NA = 1.25
Imaging Device = AxioCam 807
Camera Adapter = 1X
Depth of Focus = 1.00µm
Binning Mode = 2,2

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335-383
 Filter Emission Wavelength = 420-470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @6.0%
 Exposure Time = 18ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.25
 Imaging Device = AxioCam 807
 Camera Adapter = 1X
 Depth of Focus = 0.90µm
 Binning Mode = 2,2

Post Processing - SIM Processing

Apotome SIM Image Processing = Display Mode "Optical Sectioning" - enabled correction selected.

Normalization, Phase Correction, Fourier Filter, and Deconvolution were NOT used.

3D Image Processing

3D Volume Projection = Precise
 Appearance = Threshold 13% (488), Threshold 3% (594), Threshold13% (DAPI), Ramp 50% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 120%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, did NOT Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 182%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None

Image by E.F. Simon

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000014
Stage	Widefield SIM - Apotome
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	45

SOURCE PROVENANCE

Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	2BF827B70DAB3B85C42405F1578F79E3CD770F8BC44ED6602F8E93AC32447237
Method PDF SHA-256	B54CCD997DA79762937C4C11FE05619C883886483A8EC5F225A54A8BBBA463E5

ORIGINAL IMAGE EVIDENCE / CONFOCAL



EVIDENCE TIER	Confocal
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	44

STRUCTURED ACQUISITION METADATA / CONFOCAL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	70 Slices (4.41µm)
Channels	2
Scaling (Per Pixel)	0.071µm X 0.071µm X 0.060µm
Image Size (Pixels)	1214 X 1456
Image Size (Scaled)	85.71µm X 102.80µm
Bit Depth	16 Bit
Image Center Position	X: 75.13mm, Y: 49.74mm
Stage Position	X: 75.13mm, Y: 49.74mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Contrast Method	Fluorescence
Pinhole	1.00AU / 45µm 1.00AU / 54µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	488nm @0.2% 561nm @0.2%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	0.70
Rotation	0°
Sampling	1.00
Pixel Time	0.51µs
Frame Time	10.15s
LSM Scan Speed	7
Scan Direction	Bidirectional
Line Step	1
Excitation Wavelength	493 590
Emission Wavelength	517 618
Effective NA	1.4
Detection Wavelength	496 - 575 590 - 700
Imaging Device	GaAsp-PMT 1 GaAsp-PMT 2
Detector Type	GaAsp-PMT
Detector Gain	700 V

FIELD	SOURCE-DOCUMENTED VALUE
Detector Offset	0
Detector Digital Gain	1.0
3D Maximum Projection	Precise
Appearance	Threshold 1% (488), Threshold 2% (594), Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness100%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000048

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Mitochondria Antibody (MA5-12017), followed, and was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 594 secondary antibody was incubated for 1 hour at room temperature at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 70 Slices (4.41µm)

Channels = 2

Scaling (Per Pixel) = 0.071µm X 0.071µm X 0.060µm

Image Size (Pixels) = 1214 X 1456

Image Size (Scaled) = 85.71µm X 102.80µm

Bit Depth = 16 Bit

Image Center Position = X: 75.13mm, Y: 49.74mm

Stage Position = X: 75.13mm, Y: 49.74mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

488 -

Reflector = None

Contrast Method = Fluorescence

Pinhole = 1.00AU / 45µm

LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)

Laser Wavelength = 488nm @0.2%

Laser Blanking = Enabled

Scan Mode = Frame

Scan Zoom = 0.70

Rotation = 0°

Sampling = 1.00

Pixel Time = 0.51µs

Frame Time = 10.15s

LSM Scan Speed = 7

Scan Direction = Bidirectional

Line Step = 1

Averaging - 4

Excitation Wavelength = 493

Emission Wavelength = 517

Effective NA = 1.4

Detection Wavelength = 496 - 575

Imaging Device = GaAsp-PMT 1

Detector Type = GaAsp-PMT

Detector Gain = 700 V

Detector Offset = 0

Detector Digital Gain = 1.0

594 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00AU / 54µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 561nm @0.2%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 0.70
 Rotation = 0°
 Sampling = 1.00
 Pixel Time = 0.51µs
 Frame Time = 10.15s
 LSM Scan Speed = 7
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging - 4
 Excitation Wavelength = 590
 Emission Wavelength = 618
 Effective NA = 1.4
 Detection Wavelength = 590 - 700
 Imaging Device = GaAsP-PMT 2
 Detector Type = GaAsP-PMT
 Detector Gain = 700 V
 Detector Offset = 0
 Detector Digital Gain = 1.0

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 1% (488), Threshold 2% (594), Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness100%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None

Image by E.F. Simon

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000014
Stage	Confocal

SOURCE PROVENANCE

Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	44
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	2B8986E8AFA4388C0E451363639086347A224BD46074E03565F860F9B1117775
Method PDF SHA-256	764BFB27373ADB6284CE85E97296A663BAD845E8FDFAA26FEE710E0BBCEAD20D

ORIGINAL IMAGE EVIDENCE / AIRYSCAN



EVIDENCE TIER	Airyscan
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488
METADATA VALUES	46

STRUCTURED ACQUISITION METADATA / AIRYSCAN

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	49 Slices (2.4µm)
Channels	2
Scaling (Per Pixel)	0.03529 µm X 0.03529 µm X 0.05000 µm
Image Size (Pixels)	1262 X 670
Image Size (Scaled)	44.54µm X 23.65µm
Bit Depth	16 Bit
Image Center Position	X: 77.38mm, Y: 50.72mm
Stage Position	X: 77.38mm, Y: 50.72mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Contrast Method	Fluorescence
Pinhole	5.00 AU / 219µm 5.00 AU / 181µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	488nm @0.2% 405nm @1.50%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	1.6
Rotation	101.91°
Sampling	2.00
Pixel Time	0.59µs
Frame Time	4.44s
LSM Scan Speed	7
Scan Direction	Bidirectional
Line Step	1
Averaging	2
Excitation Wavelength	493 353
Emission Wavelength	517 465
Effective NA	1.4
Detection Wavelength	480-560 400-460
Imaging Device	Airyscan
Detector Type	GaAsp-PMT

FIELD	SOURCE-DOCUMENTED VALUE
Detector Gain	850 V
Detector Offset	0
Detector Digital Gain	1.0
AiryScan Mode	Super Resolution 7.9 (3D, Auto) Super Resolution 7.6 (3D, Auto)
3D Maximum Projection	Precise
Appearance	Threshold 3%, (488), Threshold 1% (DAPI), Ramp 0% (488), Ramp 0% (DAPI), Maximum 100% all channels
Background	Black
Light	Brightness 116%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 90%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed three times in PBS, and Identifyn 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed five times in PBS and mounted using ProLong Diamond Antifade Mountant with DAPI (cat#).

Microscope

Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 49 Slices (2.4 μ m)

Channels = 2
Scaling (Per Pixel) = 35.29nm X 35.29nm X 50nm
Image Size (Pixels) = 1262 X 670
Image Size (Scaled) = 44.54 μ m X 23.65 μ m
Bit Depth = 16 Bit
Image Center Position = X: 77.38mm, Y: 50.72mm
Stage Position = X: 77.38mm, Y: 50.72mm
ROI Center Offset = X: 0.00 μ m, Y: 0.00 μ m

488 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 5.00 AU / 219µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 488nm @0.2%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.6
 Rotation = 101.91°
 Sampling = 2.00
 Pixel Time = 0.59µs
 Frame Time = 4.44s
 LSM Scan Speed = 7
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 2
 Excitation Wavelength = 493
 Emission Wavelength = 517
 Effective NA = 1.4
 Detection Wavelength = 480-560
 Imaging Device = Airyscan
 Detector Type = GaAsP-PMT
 Detector Gain = 850 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 AiryScan Mode = Super Resolution 7.9 (3D, Auto)

DAPI -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 5.00 AU / 181µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 405nm @1.50%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.6
 Rotation = 101.91°
 Sampling = 2.00
 Pixel Time = 0.59µs
 Frame Time = 4.44s
 LSM Scan Speed = 7
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 2
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Detection Wavelength = 400-460
 Imaging Device = Airyscan
 Detector Type = GaAsP-PMT
 Detector Gain = 850 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 AiryScan Mode = Super Resolution 7.6 (3D, Auto)

Post Processing - AiryScan 3D Processing

Display Mode "SR"
 Super Resolution Auto Filter Selected
 Process 3D Current Image
 Created

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 3%, (488), Threshold 1% (DAPI), Ramp 0% (488), Ramp 0% (DAPI), Maximum 100% all channels
 Background = Black
 Light = Brightness 116%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 90%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections -**None**


Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000014
Stage	Airyscan
Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 AiryScan, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	46
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	C2B53A8AD0F0CB3A5E590EE96BA9065C36CFEF533B955177E96AF5E106F7E146
Method PDF SHA-256	59B9614BA73CA6C77AD9846134DFF129761538E9F07533E56613E6D4AD9E9C70

ORIGINAL IMAGE EVIDENCE / SIM



EVIDENCE TIER	SIM
INSTRUMENT	ZEISS Elyra
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	55

STRUCTURED ACQUISITION METADATA / SIM

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	35 Slices (1.70µm)
Channels	2 1
Scaling (Per Pixel)	0.01320 µm X 0.01320 µm X 0.05000 µm
Image Size (Pixels)	5451 X 5106
Image Size (Scaled)	71.93µm X 67.38µm
Bit Depth	16 Bit
Image Center Position	X: 60.24mm, Y: 36.77mm
Stage Position	X: 60.24mm, Y: 36.77mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflectors	SR HE LBF 405/488/561/642
Beam Splitter	405, 488, 561, 642
Filter Excitation Wavelength	200 - 405, 200 - 488, 200 - 561, 200 - 642
Filter Emission Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Contrast Method	Fluorescence
Illumination Wavelength	488 561
Channel Name	T2-Axiocam 820 -SR T1-Axiocam 820 #2-SR
Excitation Wavelength	488 561
Emission Wavelength	525 597
Effective NA	1.4
Detection Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Imaging Device	Axiocam 820 Axiocam 820#2
Camera Adapter	1X
Exposure Time	100ms 120ms
Depth of Focus	0.81µm 0.92µm
Binning Mode	2,2
Laser Wavelength	488nm @1.50% 561nm @3.0%
Frame Time	1.33s
Scan Direction	Unidirectional
Result Sampling	4
Process Sampling	2
Algorithm	SIM

FIELD	SOURCE-DOCUMENTED VALUE
SIM Settings	Standard
Auto Sharpness	On
Sharpness	9.7670
Input SNR	N/A
Regularization Weight	N/A
Iterations	N/A
Filter	N/A
Fitting	N/A
Normalization	Automatic
Experimental PSF	N/A
3D Maximum Projection	Precise
Appearance	Threshold 4% (594), Threshold 4% (488), Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 100%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 000014

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000048

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Mitochondria Antibody (MA5-12017), followed, and was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 594 secondary antibody was incubated for 1 hour at room temperature at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 35 Slices (1.70µm)

Channels = 2

Scaling (Per Pixel) = 13.20nm X 13.20nm X 50.00nm

Image Size (Pixels) = 5451 X 5106

Image Size (Scaled) = 71.93µm X 67.38µm

Bit Depth = 16 Bit

Image Center Position = X: 60.24mm, Y: 36.77mm

Stage Position = X: 60.24mm, Y: 36.77mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

488 -

Reflectors = SR HE LBF 405/488/561/642

Beam Splitter = 405, 488, 561, 642

Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642

Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800

Contrast Method = Fluorescence

Illumination Wavelength = 488

Channel Name = T2-Axiocam 820 -SR

Excitation Wavelength = 488

Emission Wavelength = 525

Effective NA = 1.4

Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800

Imaging Device = Axiocam 820

Camera Adapter = 1X

Exposure Time = 100ms

Depth of Focus = 0.81µm

Binning Mode = 2,2

Laser Wavelength = 488nm @1.50%

Frame Time = 1.33s

Scan Direction = Unidirectional

Grating 32.0µm grating

594 -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 561
 Channel Name = T1-Axiocam 820 #2-SR
 Excitation Wavelength = 561
 Emission Wavelength = 597
 Effective NA = 1.4
 Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Imaging Device = Axiocam 820#2
 Camera Adapter = 1X
 Exposure Time = 120ms
 Depth of Focus = 0.92µm
 Binning Mode = 2,2
 Laser Wavelength = 561nm @3.0%
 Frame Time = 1.33s
 Scan Direction = Unidirectional
 Grating 32.0µm grating

Post Processing - SIM Processing

Processing Dimension - 3D
 Result Sampling = 4
 Process Sampling = 2
 Channels = 1
 Algorithm = SIM
 SIM Settings = Standard
 Auto Sharpness = On
 Sharpness = 9.7670
 Input SNR = N/A
 Regularization Weight = N/A
 Iterations = N/A
 Filter = N/A
 Fitting = N/A
 Normalization = Automatic
 Experimental PSF = N/A

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 4% (594), Threshold 4% (488), Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 100%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections


None

Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000014
Stage	SIM
Instrument	ZEISS Elyra
Microscope configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Structured source metadata values	55
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	369B3B3B8ED6C8EFFCEBE4538CA66A082BCC1E91DE998EFBCF6178A2A27AB0E5
Method PDF SHA-256	77B362081E714D6A720AF2BD65C5511458CA6FACE2AAAB678480D29A149FC4A4

ORIGINAL IMAGE EVIDENCE / SIM²

EVIDENCE TIER	SIM ²
INSTRUMENT	ZEISS Elyra
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 594
METADATA VALUES	55

STRUCTURED ACQUISITION METADATA / SIM²

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	39 Slices (1.90µm)
Channels	2 1
Scaling (Per Pixel)	0.01320 µm X 0.01320 µm X 0.05000 µm
Image Size (Pixels)	2324 X 2471
Image Size (Scaled)	30.67µm X 32.61µm
Bit Depth	16 Bit
Image Center Position	X: 60.04mm, Y: 36.77mm
Stage Position	X: 60.04mm, Y: 37.35mm
ROI Center Offset	X: 0.00µm, Y: 37.35µm
Reflectors	SR HE LBF 405/488/561/642
Beam Splitter	405, 488, 561, 642
Filter Excitation Wavelength	200 - 405, 200 - 488, 200 - 561, 200 - 642
Filter Emission Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Contrast Method	Fluorescence
Illumination Wavelength	488 561
Channel Name	T2-Axiocam 820 -SR T1-Axiocam 820 #2-SR
Excitation Wavelength	488 561
Emission Wavelength	525 597
Effective NA	1.4
Detection Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Imaging Device	Axiocam 820 Axiocam 820#2
Camera Adapter	1X
Exposure Time	100ms 120ms
Depth of Focus	0.81µm 0.92µm
Binning Mode	2,2
Laser Wavelength	488nm @1.50% 561nm @3.0%
Frame Time	1.33s
Scan Direction	Unidirectional
Result Sampling	4
Process Sampling	4
Algorithm	SIM2

FIELD	SOURCE-DOCUMENTED VALUE
SIM Settings	Standard
Auto Sharpness	On
Sharpness	9.1424
Input SNR	N/A
Regularization Weight	N/A
Iterations	N/A
Filter	N/A
Fitting	N/A
Normalization	Automatic
Experimental PSF	N/A
3D Volume Projection	Precise
Appearance	Threshold 2% (594), Threshold 2% (488), Threshold 2% (DAPI), Ramp 2% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 100%, Azimuth 80° , Elongation -130° , Enabled directional Light, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Identifyn® Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG
Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human IgG (H + L)
Cat ID# - SA 0000014

Identifyn® SRM Alexa Fluor™ 594 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000048

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Mitochondria Antibody (MA5-12017), followed, and was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 594 secondary antibody was incubated for 1 hour at room temperature at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 39 Slices (1.90µm)

Channels = 2

Scaling (Per Pixel) = 13.20nm X 13.20nm X 50.00nm

Image Size (Pixels) = 2324 X 2471

Image Size (Scaled) = 30.67µm X 32.61µm

Bit Depth = 16 Bit

Image Center Position = X: 60.04mm, Y: 36.77mm

Stage Position = X: 60.04mm, Y: 37.35mm

ROI Center Offset = X: 0.00µm, Y: 37.35µm

488 -

Reflectors = SR HE LBF 405/488/561/642

Beam Splitter = 405, 488, 561, 642

Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642

Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800

Contrast Method = Fluorescence

Illumination Wavelength = 488

Channel Name = T2-Axiocam 820 -SR

Excitation Wavelength = 488

Emission Wavelength = 525

Effective NA = 1.4

Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800

Imaging Device = Axiocam 820

Camera Adapter = 1X

Exposure Time = 100ms

Depth of Focus = 0.81µm

Binning Mode = 2,2

Laser Wavelength = 488nm @1.50%

Frame Time = 1.33s

Scan Direction = Unidirectional

Grating 32.0µm grating

594 -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 561
 Channel Name = T1-Axiocam 820 #2-SR
 Excitation Wavelength = 561
 Emission Wavelength = 597
 Effective NA = 1.4
 Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Imaging Device = Axiocam 820#2
 Camera Adapter = 1X
 Exposure Time = 120ms
 Depth of Focus = 0.92µm
 Binning Mode = 2,2
 Laser Wavelength = 561nm @3.0%
 Frame Time = 1.33s
 Scan Direction = Unidirectional
 Grating 32.0µm grating

Post Processing - SIM2 Processing

Processing Dimension - 3D
 Result Sampling = 4
 Process Sampling = 4
 Channels = 1
 Algorithm = SIM2
 SIM Settings = Standard
 Auto Sharpness = On
 Sharpness = 9.1424
 Input SNR = N/A
 Regularization Weight = N/A
 Iterations = N/A
 Filter = N/A
 Fitting = N/A
 Normalization = Automatic
 Experimental PSF = N/A

3D Image Processing

3D Volume Projection = Precise
 Appearance = Threshold 2% (594), Threshold 2% (488), Threshold 2% (DAPI), Ramp 2% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 100%, Azimuth 80° , Elongation -130° , Enabled directional Light, Enabled Tone Mapping
 Projection = View Angle 35° , Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None


Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000014
Stage	SIM²
Instrument	ZEISS Elyra
Microscope configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Structured source metadata values	55
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	134ECA4381DCFF3389A5750472D103C50243A2D3B37DF70E8C585F414A0B431A
Method PDF SHA-256	1F0FBF4C0F2F25BC7D8D5349BDB39EAE0EE6779C2B713F7ABA5D0A5A7FC664BA

ORIGINAL IMAGE EVIDENCE / SINGLE-MOLECULE STORM



EVIDENCE TIER	Single-molecule STORM
INSTRUMENT	Bruker Vutara VXL
CELL MODEL	HeLa
DETECTED DYES	488
METADATA VALUES	71

STRUCTURED ACQUISITION METADATA / SINGLE-MOLECULE STORM

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:
Objective	Olympus 60X/1.30 silicon oil objective
Number of Probes	1
Number of Simultaneous Probes	1
Number of Reference Probes	1
Camera	Both Cameras
Color Channel	Both Channels
Probe Capture Mode	Interleaved
Z -Stack Capture Mode	Interleaved
Multiple Location Capture	Not selected
Post Record Reference	Not Selected
Fluidics	Not Selected
Autofocus Drift Correction	On
Biplane Module	635nm Dichroic
Objective	Silicon Selected
Stage Insert	Rectangular Selected
Course Focus Position	8-Well Chamber Selected
Dichroic	SuperRes
PSF	(Red, Orange, Green)
Heater	Current: 26
Recording Vertical Field of View	100%
Widefield Camera Exposure Time	100ms 20ms
Super Resolution Camera Exposure Time	20ms
Widefield Camera Binning	1
Per-Probe Exposure Time	Selected
Predefined Probe	First probe: 488 nm (635 nm Dichroic); First reference probe: 488 nm (635 nm Dichroic)
Photoactivation/Imaging Laser	0% Selected
Exposure Time	20ms Selected
Record Workflow	Selected Options
View Mode	Widefield 200µm
Capture Mode	Live
Piezo Current	179.8 µm

FIELD	SOURCE-DOCUMENTED VALUE
Widefield 488nm	0.1X Neutral Density Filter: Selected, @0.2% Laser Input
Reference Probe 1 Laser control (488 nm (635 nm Dichroic))	Not Selected
Widefield camera Exposure Time	100ms
Widefield Laser	"Widefield 488nm" was deselected, and "calibrate" under Autofocus was selected.
Current Intensity	4.5
Calibration Value	-5084 nm/unit
Autofocus Mode	Pre-Capture
SuperRes	50µm is Selected
Piezo Record	Begin: 180; End: 180
SuperRes 488nm	0.1X Neutral Density Filter: Selected, @0.02% Laser Input
Record Workflow - Record Tab	Capture Parameters
Frames	Probe 1: 15000
Z Positions	1 using Steps: 0.2µm (200nm)
Cycles	1
Time Points	1
Pause Between Cycles	Not Selected
Automatic Photoactivation Laser Control	Not Selected
Laser Power	15%
Cutout Width	16px
Max Frame Accumulation	1 frame
Max Accumulation Offset	1px
Fiducial at Max Accumulation	Not selected
Max Particles Per Frame	No Max
Background SLPE Removal	Not Selected
Cutout/Background Removal	Not Selected
Preview Cutout Positions	Selected
Probe1	BG threshold: 10
Sizing Mode	Radial Precision
Particle size	5nm
Color by	Constant
Color Table	Single
Opacity Mode	Constant; Opacity 1: 05
Front-to-back Attenuation	Not Selected
Method	Particle (direct)
Time Intervals	5 Intervals

FIELD	SOURCE-DOCUMENTED VALUE
Pixel Size	20nm
Smoothing	8.00
PSF Z offset	-450 to +450
Radial precision	30
Axial precision	70

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn Primary Antibody

Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG

Cat ID# - PA-RR 000004

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human, IgG (H + L)

Cat ID# - SA 000014

Gamma H2AX (γ H2AX) is a variant of the histone H2A protein that plays a crucial role in the cellular response to DNA double-strand breaks (DSBs). Its primary function is as a marker of DNA damage, particularly in the context of DSBs. When DSBs occur in chromatin, γ H2AX is rapidly phosphorylated explicitly at the serine 139 residue (referred to as Ser139 phosphorylation or γ -phosphorylation) near the site of the break. This phosphorylation event forms γ H2AX foci, visualized via immunofluorescence microscopy techniques as an aggregate of fluorescently labeled protein at the site of breaks.

Method

HeLa cells were grown in a μ -Slide 8 Well high Glass-Bottom Ibidi chamber. DNA damage was induced via treatment with 2 μ M Etoposide-spiked complete growth media for 6 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human γ H2Ax Recombinant Monoclonal, Human Fc, IgG, was incubated for 1 hour at room temperature at a dilution of 1:100. Cells were washed 3X in Identifyn's™ proprietary Wash Buffer and Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:10. Cells were washed 5X in Identifyn's™ proprietary Wash Buffer followed by the addition of Identifyn's™ proprietary STORM Buffer to the sample chamber enabling image acquisition.

Microscope

Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:

Calibration -

Calibration follows Identifyn's Calibration Standard Operating Procedure, which can be found here.

<https://identifyn.com/storm-calibration>

Image Acquisition -

SRX Software - Under the 'Single Molecule Localization' tab, 'Record' is selected and contains the following workflow tabs and parameters.

Configuration Tab - Capture Configuration

Number of Probes: 1
 Number of Simultaneous Probes: 1
 Number of Reference Probes: 1
 Camera: Both Cameras
 Color Channel: Both Channels
 Probe Capture Mode: Interleaved
 Z -Stack Capture Mode: Interleaved
 Multiple Location Capture: Not selected
 Post Record Reference: Not Selected
 Fluidics: Not Selected
 Autofocus Drift Correction: On

Microscope configuration

Biplane Module: 635nm Dichroic
 Objective: Silicon Selected
 Stage Insert: Rectangular Selected
 Course Focus Position: 8-Well Chamber Selected
 Dichroic: SuperRes
 PSF: (Red, Orange, Green)
 Heater: Current: 26

Camera Configuration

Recording Vertical Field of View: 100%
 Widefield Camera Exposure Time: 100ms
 Super Resolution Camera Exposure Time: 20ms
 Widefield Camera Binning: 1
 Per-Probe Exposure Time: Selected

Probe Configuration

Predefined Probe: First probe: 488 nm (635 nm Dichroic); First reference probe: 488 nm (635 nm Dichroic)
 Photoactivation/Imaging Laser: 0% Selected
 Exposure Time: 20ms Selected

Experiment Configuration

Record Workflow: Selected Options
 View Mode: Widefield 200µm
 Capture Mode: Live

Stage and Piezo Selection

Examination of the Sample

Cell(s) or Cellular Structure Localization - The red-bordered box identifies the area of the SuperRes view and defines a suitable cell(s) or cellular structure. The cell(s) or structures are optimally focused.

Piezo Current: 179.8 μ m

Probe 1 Laser control (488 nm (635 nm Dichroic)

Widefield 488nm: 0.1X Neutral Density Filter: Selected, @0.2% Laser Input
Reference Probe 1 Laser control (488 nm (635 nm Dichroic): Not Selected

Advanced Tools Option

Widefield camera Exposure Time: 100ms
Default values are maintained unless otherwise noted.

Widefield Laser: "Widefield 488nm" was deselected, and "calibrate" under Autofocus was selected.

Autofocus - Calibration Values

Current Position" -0.979
Current Intensity: 4.5
Calibration Value: -5084 nm/unit
Autofocus Mode: Pre-Capture

View Mode

SuperRes: 50 μ m is Selected

Capture Mode

Live

Stage and Piezo Selection

Cell(s) or Cellular Structure Localization - Positioning of the target Cell or Subsection of the Target Cell is optimized and optimally focused.

Piezo Record: Begin: 180; End: 180

Probe 1 Laser control (488 nm (635 nm Dichroic)

SuperRes 488nm: 0.1X Neutral Density Filter: Selected, @0.02% Laser Input

Reference Probe 1 Laser control (488 nm (635 nm Dichroic)

Not Selected

Advanced Tools

Widefield Camera Exposure Time: 20ms
Default values are maintained unless otherwise noted.

Record Workflow - Record Tab: Capture Parameters

Frames: Probe 1: 15000

Z Positions: 1 using Steps: 0.2µm (200nm)

Cycles: 1

Time Points: 1

Pause Between Cycles: Not Selected

Automatic Photoactivation Laser Control: Not Selected

Probe 1 Laser control (488 nm (635 nm Dichroic)

Laser Power: 15%

Expert Configuration Options - The Localization Tab

Beyond the values shown, all other parameters were set to their default values.

Localization

Cutout Width: 16px

Max Frame Accumulation: 1 frame

Max Accumulation Offset: 1px

Fiducial at Max Accumulation: Not selected

Max Particles Per Frame: No Max

Background SLPE Removal: Not Selected

Cutout/Background Removal: Not Selected

Preview Cutout Positions: Selected

Record Workflow - Localization Tab

Data Collection

The folder is created automatically with the localization file for this image acquisition.

Recording Summary

All parameters were carried over from previous tabs and are kept except:

Probe1: BG threshold: 10

Region of Interest -

Not selected

Display

Was deployed to inspect if the BG threshold applied chooses localizations in an expected manner

Record Workflow - View Tab

Region of Interest

Default values

Visualization Parameters

Point Splatting

Sizing Mode: Radial Precision

Particle size: 5nm

Color by: Constant

Color Table: Single

Opacity Mode: Constant; Opacity 1: 05

Front-to-back Attenuation: Not Selected

Post Acquisition analysis - View Tab

Drift Correction - Computed with the following parameters:

Method: Particle (direct)

Time Intervals: 5 Intervals

Pixel Size: 20nm

Smoothing: 8.00

Advanced Particle Filters - Applied to the drift-corrected data with the following parameters

PSF Z offset: -450 to +450

Radial precision: 30

Axial precision: 70

The other parameters were not applied.

Statistical analysis - N/A

Image by S. Thakur

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000014
Stage	Single-molecule STORM
Instrument	Bruker Vutara VXL
Microscope configuration	Bruker Vutara VXL Biplane Single Molecule Microscope, using an Olympus 60X/1.30 silicon oil objective, was used with the following parameters:
Structured source metadata values	71
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa

SOURCE PROVENANCE

Image SHA-256	315D39CE59CF71FD49803FA1A8E14A69BC5B9AB46DDD55556368B97E536BFEEF
Method PDF SHA-256	D10EED14E1AE7107F67E84CBD0CB0E1572E667E339E58A49D9352BF0D8A6B05F

ORIGINAL IMAGE EVIDENCE / CONTROL



EVIDENCE TIER	Control
INSTRUMENT	ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488
METADATA VALUES	30

STRUCTURED ACQUISITION METADATA / CONTROL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Channels	2
Scaling (Per Pixel)	0.143µm X 0.143µm
Image Size (Pixels)	1422 X 960
Image Size (Scaled)	206.00µm X 137.14µm
Bit Depth	14 Bit
Image Center Position	X: 52.34mm, Y: 50.77mm
ROI Center Offset	X: 52.34µm, Y: 50.77µm
Reflector	38 HE eGFP 49 DAPI
Beam Splitter	495 395
Filter Excitation Wavelength	450 - 490 335-383
Filter Emission Wavelength	500 -550 420-470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @25% X-Cite Xylis Lamp Intensity @10%
Exposure Time	200ms 30ms
Excitation Wavelength	493 353
Emission Wavelength	517 465
Effective NA	1.25
Imaging Device	Axiocam 807 mono
Camera Adapter	1X
Depth of Focus	1.0µm 0.90µm
Binning Mode	2,2

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 488 Affinity Purified Goat Anti-Human, IgG (H + L)

Cat ID# - SA 000014

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocking with Thermo Fisher Fish Serum Blocker (cat#37527X3). In the absence of a primary antibody, the Identifyn® 488 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. The cells were washed five times in PBS and then mounted using ProLong™ Diamond Antifade Mountant with DAPI (cat# P36962).

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.143µm X 0.143µm

Image Size (Pixels) = 1422 X 960

Image Size (Scaled) = 206.00µm X 137.14µm

Bit Depth = 14 Bit

Image Center Position = X: 52.34mm, Y: 50.77mm

ROI Center Offset = X: 52.34µm, Y: 50.77µm

488 -

Reflector = 38 HE eGFP

Beam Splitter = 495

Filter Excitation Wavelength = 450 - 490

Filter Emission Wavelength = 500 -550

Contrast Method = Fluorescence

Light source = X-Cite Xylis Lamp Intensity @25%

Exposure Time = 200ms

Excitation Wavelength = 493

Emission Wavelength = 517

Effective NA = 1.25

Imaging Device = Axiocam 807 mono

Camera Adapter = 1X

Depth of Focus = 1.0µm

Binning Mode = 2,2

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335-383
 Filter Emission Wavelength = 420-470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @10%
 Exposure Time = 30ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 0.90µm
 Binning Mode = 2,2

Post Processing -

None

Image Corrections

None

Control Summary -

In the absence of a primary antibody, Identifyn Standards for Optical Microscopy Methods™ were applied using standard staining methodology and image acquisition settings using the Zeiss Axio Imager.M1 Widefield. With no image correction, the image demonstrates that the secondary antibody, minus a primary target, had no off-target binding and no significant presence of free dye within the sample.

The data suggest the secondary antibody is highly specific and free of unconjugated (free) dye.

Image by J.K. Bennett

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000014
Stage	Control
Instrument	ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	30
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa

SOURCE PROVENANCE

Image SHA-256	6920E91184FFC0DACEB578D375920D693F37E35740C44EC1EE2331D13BC35C8D
Method PDF SHA-256	660CDF2237E1A1E877DE9BE4F4A8AFDACA9DE11D7C8126BE3AAB92B0FCC5F24E