

IDENTIFYN RESOLUTION STEPDOWN

SECONDARY ANTIBODY MICROSCOPY EVIDENCE ASSESSMENT

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse, F(ab')₂ (H + L)

WESTERN BLOT -> WIDEFIELD -> WIDEFIELD SIM / APOTOME -> CONFOCAL -> AIRYSCAN
SIM -> SIM²

7


ORDERED EVIDENCE TIERS

1

SUPPORTING CONTROL

PENDING

SCIENTIST REVIEW



REPRESENTATIVE HIGHEST-AVAILABLE VIEW / SIM²

SA-000011 | PRODUCT-LEVEL EVIDENCE DOSSIER

SOURCE-GROUNDED MICROSCOPY + BIOLOGICAL REVIEW / IDENTIFYN EVIDENCE PRESERVED

CRO CAPABILITY DEMONSTRATED

Identifyn performed the experimental and imaging work documented below. BENNETT converts that operating history into a governed, source-bounded scientific record.

- Source-documented sample preparation and antibody/reagent workflow
- Western blot imaging
- Widefield fluorescence microscopy
- Widefield structured illumination / Apotome optical sectioning
- Confocal microscopy
- Airyscan acquisition and reconstruction workflow
- Structured illumination microscopy
- SIM² reconstruction workflow
- Preserved control experiment
- Z-stack / 3D acquisition
- Multi-channel imaging

SCIENTIFIC QUESTION

What source-bounded experimental and microscopy evidence is preserved for Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse, F(ab')₂ (H + L), and what is the strongest interpretation allowed by the available modalities and controls?

EVIDENCE AVAILABLE

Western blot -> Widefield -> Widefield SIM — Apotome -> Confocal -> Airyscan -> SIM -> SIM² -> Control

BENNETT ASSESSMENT

The preserved SA-000011 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

BENNETT / MICROSCOPY AND BIOLOGICAL EVIDENCE REVIEW

SOURCE-GROUNDED RETROSPECTIVE REVIEW / PENDING SCIENTIST REVIEW

SOURCE STATE	SOURCE HASH VERIFIED / EVENT RECORDED
BENNETT EVIDENCE REVIEW	COMPLETED / SOURCE-BOUNDED
NATIVE IMAGE REPROCESSING	NOT PERFORMED
LITERATURE SOURCES	INDEPENDENTLY VERIFIED
SCIENTIST REVIEW	PENDING
PUBLICATION	NOT AUTHORIZED

OVERALL CALL / NIR-BOUNDED PARTIAL STEPDOWN

The preserved SA-000011 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

SETTINGS ASSESSMENT

The record preserves 7 ordered evidence tier(s). Structured source metadata values are reported by stage and remain tied to the original method objects.

BIOLOGICAL SUPPORT

The record supports retrospective, source-bounded microscopy review. Target identity, specificity, treatment effect, binding, interaction and mechanism are not inferred from presentation images alone.

CONTROL ASSESSMENT

A true preserved control record is present and appears last.

CLAIM CEILING

This draft supports historical capability documentation, source registration, modality ordering, settings review and bounded interpretation. It does not independently validate antibody specificity, quantify a population response, establish binding, interaction, causation or mechanism, or authorize publication.

PROCESSING DISCLOSURE

Retrospective BENNETT architectural evaluation of preserved Identifyn source evidence. BENNETT did not perform native-file reprocessing, new deterministic measurement or the historical image post-processing represented in this record.

REAGENTS AND ACQUIRED CHANNELS

Preparation reagents are reported separately from channels actually documented in the acquisition settings. Fluorophores mentioned in staining are not automatically treated as acquired channels.

EVIDENCE OBJECT	REAGENTS / FLUOROPHORES PRESENT	CHANNELS ACTUALLY ACQUIRED
Western blot	405/DAPI	Not deterministically stated in the acquisition block
Widefield	405/DAPI; 555	2; AF405-49027; AF532-49014; 372 - 408; 515 - 545; 417 - 445; 555 - 595; 401
Widefield SIM — Apotome	405/DAPI; 555	2; AF405-49027; AF532-49014; 372 - 408; 515 - 545; 417 - 445; 555 - 595; 401
Confocal	405/DAPI; 488; 750	1; None; 405nm @0.9%; 401; 422
Airyscan	405/DAPI; 488	1; None; 405nm @1.0%; 401; 422
SIM	405/DAPI; 488	1; SR HE LBF 405/488/561/642; 200 - 405, 200 - 488, 200 - 561, 200 - 642; 419 - 470, 503 - 546, 576 - 617, 657 - 800; T2-Axiocam 820#2-SR; 405; 445; 405nm @3.0%
SIM ²	405/DAPI; 488	1; SR HE LBF 405/488/561/642; 200 - 405, 200 - 488, 200 - 561, 200 - 642; 419 - 470, 503 - 546, 576 - 617, 657 - 800; T2-Axiocam 820#2-SR; 405; 445; 405nm @3.0%
Control	405/DAPI; 647	2; AF405-49027; 50 Cy5; 372 - 408; 625 - 655; 417 - 445; 665 - 715; 401

MODALITY ASSESSMENT / PART 1 OF 8

Western blot / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: Azure Sapphire FL. Objective: not explicitly stated. Platform: Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s). Structured source metadata values: 11.

VISIBLE OBSERVATION

A preserved presentation image is available for Western blot. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Western blot, documents platform context as Azure Sapphire FL, and records detected dye/channel descriptors as 405/DAPI.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Band position and displayed intensity do not independently establish analytical specificity, target identity, linearity, or treatment effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Western blot evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 2 OF 8

Widefield / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Imager.M1. Objective: Plan-Apochromat 63x/1.40 Oil M27. Platform: Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam 705, and X-Cite Xylis Lamp was used with the following parameters for each dye:. Acquisition: Channels: 2; Scaling (Per Pixel): 0.102µm X 0.102µm; Image size (Pixels): 1216 X 762; Image Size (Pixels): 1216 X 762; Bit Depth: 14 Bit. Detector/camera: Imaging Device: Axiocam 705 mono; Axiocam 807 mono. Timing: Exposure Time: 500ms; 60ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @18%; X-Cite Xylis Lamp Intensity @8.0%. Optical/scan: Effective NA: 1.4. Structured source metadata values: 31.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield, documents platform context as ZEISS Axio Imager.M1, and records detected dye/channel descriptors as 405/DAPI; 555.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

A presentation image supports gross localization and source-bounded co-occurrence, not direct binding, molecular colocalization, or native quantitative measurement.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 3 OF 8

Widefield SIM — Apotome / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylys Lamp was used with the following parameters for each dye: Acquisition: Z-Stack: 78 Slices (7.7µm); Channels: 2; Scaling (Per Pixel): 0.143µm X 0.143µm X 0.100µm; Image size (Pixels): 887 X 963; Image Size (Pixels): 887 X 963; Bit Depth: 16 Bit. Detector/camera: Imaging Device: AxioCam 807 mono. Timing: Exposure Time: 700ms; 80ms. Illumination: Light source: X-Cite Xylys Lamp Intensity @25%; X-Cite Xylys Lamp Intensity @8.0%. Optical/scan: Effective NA: 1.25. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 43.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield SIM — Apotome. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield SIM — Apotome, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 555.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Improved optical sectioning does not by itself establish reconstructed super-resolution or molecular interaction.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield SIM — Apotome evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 4 OF 8

Confocal / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 44 Slices (3.01µm); Channels: 1; Scaling (Per Pixel): 0.071µm X 0.071µm x 0.070µm; Image size (Pixels): 1330 X 1554; Image Size (Pixels): 1330 X 1554; Bit Depth: 16 Bit. Detector/camera: Imaging Device: GaAsP-Pmt1; Detector Type: GaAsP-PMT. Timing: Pixel Time: 0.33µs; Frame Time: 3.95s. Illumination: Laser Wavelength: 405nm @0.9%. Optical/scan: Pinhole: 1.00 AU / 36µm; Scan Zoom: 0.90; LSM Scan Speed: 9; Scan Direction: Bidirectional; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 59%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 37.

VISIBLE OBSERVATION

A preserved presentation image is available for Confocal. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Confocal, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488; 750.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Optical sectioning and source-bounded 3D presentation do not establish binding, mechanism, or population-level effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Confocal evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 5 OF 8

Airyscan / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 900. Objective: with a Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:. Acquisition: Z-Stack: 49 Slices (2.40µm); Channels: 1; Scaling (Per Pixel): 35.30nm x 35.30nm x 50.00nm; Image size (Pixels): 2389 X 1747; Image Size (Pixels): 2389 X 1747; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Airyscan; Detector Type: GaAsP-PMT. Timing: Pixel Time: 0.73µs; Frame Time: 14.22s. Illumination: Laser Wavelength: 405nm @1.0%. Optical/scan: Pinhole: 5.00 AU / 183µm; Scan Zoom: 1.0; LSM Scan Speed: 6; Scan Direction: Bidirectional; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 38.

VISIBLE OBSERVATION

A preserved presentation image is available for Airyscan. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Airyscan, documents platform context as ZEISS LSM 900, and records detected dye/channel descriptors as 405/DAPI; 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Displayed feature separation remains reconstruction dependent; no unsupported numerical resolution or molecular architecture claim is permitted.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Airyscan evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 6 OF 8

SIM / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:. Acquisition: Z-Stack: 94 Slices (4.65µm); Channels: 1; Scaling (Per Pixel): 16.89nm X 16.89nm X 50.00nm; Image size (Pixels): 5056 X 5056; Image Size (Pixels): 5056 X 5056; Bit Depth: 16 Bit. Detector/camera: Imaging Device: AxioCam 820#2. Timing: Exposure Time: 100ms; Frame Time: 676.00ms. Illumination: Laser Wavelength: 405nm @3.0%; Illumination Wavelength: 405. Optical/scan: Scan Direction: Unidirectional; Effective NA: 1.4. Processing: Algorithm: SIM; Process Sampling: N/A; Result Sampling: 4; Projection: View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 42.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM, documents platform context as ZEISS Elyra, and records detected dye/channel descriptors as 405/DAPI; 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Reconstruction-dependent structural separation requires native reconstruction QC before numerical resolution or molecular dimensions can be claimed.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 7 OF 8

SIM² / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Elyra. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:. Acquisition: Z-Stack: 62 Slices (3.05µm); Channels: 1; Scaling (Per Pixel): 21.11nm X 21.11nm X 50.00nm; Image size (Pixels): 3525 X 3946; Image Size (Pixels): 3525 X 3946; Bit Depth: 16 Bit. Detector/camera: Imaging Device: AxioCam 820#2. Timing: Exposure Time: 100ms; Frame Time: 676.00ms. Illumination: Laser Wavelength: 405nm @3.0%; Illumination Wavelength: 405. Optical/scan: Scan Direction: Unidirectional; Effective NA: 1.4. Processing: Algorithm: SIM2; Process Sampling: 4; Result Sampling: 4; Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 45.

VISIBLE OBSERVATION

A preserved presentation image is available for SIM². BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as SIM², documents platform context as ZEISS Elyra, and records detected dye/channel descriptors as 405/DAPI; 488.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy guidance that advanced-resolution outputs remain dependent on acquisition, reconstruction or localization quality control.

INTERPRETIVE LIMIT

Further reconstruction-dependent segmentation does not establish molecular dimensions, structure identity, or mechanism without independent support.

CLAIM CEILING

This tier supports a source-bounded statement that the historical SIM² evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 8 OF 8

Control / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye: Acquisition: Channels: 2; Scaling (Per Pixel): 0.143µm X 0.143µm; Image size (Pixels): 1192 X 947; Image Size (Pixels): 1192 X 947; Bit Depth: 14 Bit. Detector/camera: Imaging Device: AxioCam 807 mono. Timing: Exposure Time: 650ms; 200ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @25%. Optical/scan: Effective NA: 1.25. Structured source metadata values: 29.

VISIBLE OBSERVATION

A preserved presentation image is available for Control. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Control, documents platform context as ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 647.

LITERATURE CORRELATION

Consistent with peer-reviewed antibody-control guidance that a no-primary or related control bounds background but is not a complete specificity system.

INTERPRETIVE LIMIT

A control bounds the recorded comparison only; it does not establish universal specificity, zero background, zero cross-reactivity, or absence of free dye.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Control evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

CROSS-MODALITY SYNTHESIS

The preserved SA-000011 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal; Airyscan; SIM; SIM². The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

Different acquisitions are not treated as a quantitative intensity ladder. Any continuity statement remains qualitative and source bounded.

BENNETT LITERATURE CORRELATION — SOURCES INDEPENDENTLY VERIFIED

BENNETT uses independently verified primary scientific sources to test whether the interpretation is scientifically consistent and properly bounded. Literature correlation does not retroactively validate the Identifyn dataset or replace scientist review.

- Antibody and microscopy evidence require orthogonal controls and modality-appropriate interpretation; presentation images alone do not establish analytical specificity.
- Advanced-resolution presentation depends on acquisition, reconstruction or localization quality control before numerical resolution, molecular dimensions or cluster identity can be claimed.

REFERENCES

1. Uhlen M, et al. A proposal for validation of antibodies. Nat Methods. 2016;13:823-827. doi:10.1038/nmeth.3995.
2. Gustafsson MGL. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. J Microsc. 2000;198:82-87. doi:10.1046/j.1365-2818.2000.00710.x.
3. Sage D, et al. Quantitative evaluation of software packages for single-molecule localization microscopy. Nat Methods. 2015;12:717-724. doi:10.1038/nmeth.3442.

RELEASE BOUNDARY

This assessment remains a draft pending scientist review. No publication is authorized. Any future native-file analysis must enter BENNETT as a new evidence snapshot with routed engine authority, native-data QA, methods, limitations, and an independently reviewed claim ceiling.

SOURCE RECONCILIATION / QA / PROVENANCE

SOURCE STATUS - PRESERVED / AUTO-RECONCILED

Historical Identifyn source material was subjected to BENNETT retrospective QA. Original source state and correction history remain preserved in provenance.

HIGH CONFIDENCE - AUTO-RECONCILED

Airyscan / Scaling (Per Pixel): 35.30nm x 35.30nm x 50.00nm -> 0.03530 μm X 0.03530 μm X 0.05000 μm . Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=2389 X 1747; Image Size (Scaled)=84.33 μm X 61.67 μm ; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.00%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM / Scaling (Per Pixel): 16.89nm X 16.89nm X 50.00nm -> 0.02111 μm X 0.02111 μm X 0.05000 μm . Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=5056 X 5056; Image Size (Scaled)=106.75 μm X 106.75 μm ; PASS - INTERNALLY CONSISTENT. The source magnitude/unit pair fails by approximately 1,000-fold; independent X and Y field dimensions establish the reconciled sampling. Maximum field-dimension error 0.00%.

HIGH CONFIDENCE - AUTO-RECONCILED

SIM² / Scaling (Per Pixel): 21.11nm X 21.11nm X 50.00nm -> 0.02111 μm X 0.02111 μm X 0.05000 μm . Both numerical magnitude and unit were normalized, then checked against the reported field dimensions; no unit label was changed alone. Supporting evidence: Image Size (Pixels)=3525 X 3946; Image Size (Scaled)=74.42 μm X 83.81 μm ; PASS - INTERNALLY CONSISTENT. Documented sampling reproduces the reported physical field dimensions. Maximum field-dimension error 0.61%.

BENNETT EVIDENCE REVIEW COMPLETE

INTERPRETATION TO EVIDENCE

BENNETT has completed the bounded retrospective assessment.
The following pages switch ownership to the preserved Identifyn record.



IDENTIFYN PRESERVED SOURCE EVIDENCE

ORIGINAL RECORD BEGINS

Images, source methods, acquisition settings and provenance follow in the locked
Stepdown order. Scientific wording is preserved; missing modalities are not invented.
A preserved control follows all available Stepdown tiers as supporting evidence.

PRODUCT-LEVEL STEPDOWN MAP


This dossier preserves the complete available evidence progression for SA-000011. Each tier remains tied to its original image, source method, instrument context and cryptographic identity. A preserved control is retained as supporting evidence and appears after every resolution tier.

CANONICAL STEPDOWN	EVIDENCE STAGE	INSTRUMENT	STATUS
1	Western blot	Azure Sapphire FL	PRESERVED
2	Widefield	ZEISS Axio Imager.M1	PRESERVED
3	Widefield SIM - Apotome	ZEISS Axio Observer.Z1/7 Apotome	PRESERVED
4	Confocal	ZEISS LSM 900	PRESERVED
5	Airyscan	ZEISS LSM 900	PRESERVED
6	SIM	ZEISS Elyra	PRESERVED
7	SIM ²	ZEISS Elyra	PRESERVED

INTERPRETIVE BOUNDARY

Ordering evidence by resolution tier does not make the modalities interchangeable. Each stage retains its acquisition environment and scientific limitations. This pilot organizes the record; it does not synthesize new biological claims.

ORIGINAL IMAGE EVIDENCE / WESTERN BLOT



EVIDENCE TIER	Western blot
INSTRUMENT	Azure Sapphire FL
CELL MODEL	HeLa
DETECTED DYES	405/DAPI
METADATA VALUES	11

STRUCTURED ACQUISITION METADATA / WESTERN BLOT

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Scanner	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Detection mode	Fluorescence
Laser	375nm
Emission Filter	452±45nm
Detector	PMT
Intensity	5
Pixel size	100 µm
Scan Speed	Medium
Quality	High
Focus	Membrane (0 µm in Z-axis)
Image	Identifyn® LLC

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

HeLa cells were homogenized using an ultrasonication probe and centrifuged to pellet down the cell debris. The lysate supernatant was mixed with NuPAGE™ LDS Sample Buffer (4X), heated at 80oC for 10 min, and subjected to electrophoresis using NuPAGE™ 4 to 12%, Bis-Tris, 1.0-1.5 mm, Mini Protein Gels in a Mini Gel Tank. The Thermo Fisher's PageRuler™ Prestained Protein Ladder, 10 to 180 kDa, was used as a molecular weight marker.

Post electrophoresis, the proteins were transferred to a polyvinylidene difluoride (PVDF) membrane using iBlot™ 3 Starter Kit, PVDF. The transfer membrane was blocked with Blocker™ Casein for 1 hour and washed 5X with PBS. The blocked membrane was incubated with Actin Monoclonal Antibody (mAbGEa) (MA1-744) for 2 hours, followed by 5X PBS washes. Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit anti-Mouse F(ab)₂ was incubated for 1 hour and washed 5X with PBS. The membrane was dried for 10 minutes at room temperature before imaging.

Scanner = Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

Scan Settings

Detection mode = Fluorescence

Laser = 375nm

Emission Filter = 452±45nm

Detector = PMT

Intensity = 5

Pixel size = 100 µm

Scan Speed = Medium

Quality = High

Focus = Membrane (0 µm in Z-axis)

Post Processing

None

Image: Identifyn® LLC

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000011
Stage	Western blot
Instrument	Azure Sapphire FL
Microscope configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

SOURCE PROVENANCE

Structured source metadata values	11
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	DD886194FF658A84D38B995043F42B426F81F175FEBE59FE317CE6B8FA517264
Method PDF SHA-256	A16A9CE4AD315C52E60AA186F5BFCF31F1DD481506E56DBD69AC68D240404141

ORIGINAL IMAGE EVIDENCE / WIDEFIELD



EVIDENCE TIER	Widefield
INSTRUMENT	ZEISS Axio Imager.M1
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 555
METADATA VALUES	31

STRUCTURED ACQUISITION METADATA / WIDEFIELD

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam 705, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	Plan-Apochromat 63x/1.40 Oil M27
Channels	2
Scaling (Per Pixel)	0.102µm X 0.102µm
Image size (Pixels)	1216 X 762
Image Size (Scaled)	133.18µm X 83.46µm
Bit Depth	14 Bit
Image Center Position	X: 0.00µm, Y: 0.00µm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	AF405-49027 AF532-49014
Beam Splitter	412 550
Filter Excitation Wavelength	372 - 408 515 - 545
Filter Emission Wavelength	417 - 445 555 - 595
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @18% X-Cite Xylis Lamp Intensity @8.0%
Exposure Time	500ms 60ms
Excitation Wavelength	401 553
Emission Wavelength	422 568
Effective NA	1.4
Imaging Device	Axiocam 705 mono Axiocam 807 mono
Camera Adapter	1X
Depth of Focus	0.65µm 0.88µm
Binning Mode	2,2

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Identifyn® SRM Alexa Fluor™ 555 Affinity Purified Goat Anti-Rabbit, IgG (H + L)

Cat ID# - SA 000035

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher, Fish Serum Blocker cat#37527X3. The first primary antibody, Nuclear Pore Complex Monoclonal Antibody (39C7)

, was incubated overnight at 4°C at a dilution of 1:100. Cells were washed 3X in PBS. Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody followed this, Actin Recombinant Rabbit Monoclonal Antibody (JF47-01), which was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 555 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, Axiocam 705, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.102µm X 0.102µm

Image size (Pixels) = 1216 X 762

Image Size (Scaled) = 133.18µm X 83.46µm

Bit Depth = 14 Bit

Image Center Position = X: 0.00µm, Y: 0.00µm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

405 -

Reflector = AF405-49027
 Beam Splitter = 412
 Filter Excitation Wavelength = 372 - 408
 Filter Emission Wavelength = 417 - 445
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @18%
 Exposure Time = 500ms
 Excitation Wavelength = 401
 Emission Wavelength = 422
 Effective NA = 1.4
 Imaging Device = Axiocam 705 mono
 Camera Adapter = 1X
 Depth of Focus = 0.65µm
 Binning Mode = 2,2

555 -

Reflector = AF532-49014
 Beam Splitter = 550
 Filter Excitation Wavelength = 515 - 545
 Filter Emission Wavelength = 555 - 595
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @8.0%
 Exposure Time = 60ms
 Excitation Wavelength = 553
 Emission Wavelength = 568
 Effective NA = 1.4
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 0.88µm
 Binning Mode = 2,2

Post Processing

None

Image Correction

Identifyn® SRM Alexa Fluor™ 555 was pseudo-colored in Zen Software from orange, the default color, to dark grey to provide visual contrast between the 405/555 channels.

Image by J.K. Bennett

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000011
---------	-----------

SOURCE PROVENANCE

Stage	Widefield
Instrument	ZEISS Axio Imager.M1
Microscope configuration	Zeiss Axio Imager.M1 Widefield, using a Plan-Apochromat 63x/1.40 Oil M27, AxioCam 705, and X-Cite Xylys Lamp was used with the following parameters for each dye:
Structured source metadata values	31
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	9B24C27FDA7AD736B098C6224812348358A0E50F80FB0AF03FA8DEE91CE8FF55
Method PDF SHA-256	9E70226A0CFFC0726E7D2913FF8DC6A44A4E706D9EDE1DED7876C1D7D9407E51

ORIGINAL IMAGE EVIDENCE / WIDEFIELD SIM - APOTOME



EVIDENCE TIER	Widefield SIM - Apotome
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 555
METADATA VALUES	43

STRUCTURED ACQUISITION METADATA / WIDEFIELD SIM - APOTOME

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Z-Stack	78 Slices (7.7µm)
Channels	2
Scaling (Per Pixel)	0.143µm X 0.143µm X 0.100µm
Image Size (Pixels)	887 X 963
Image Size (Scaled)	126.71µm X 137.57µm
Bit Depth	16 Bit
Image Center Position	X: 91.55mm, Y: 45.15mm
Stage Position	91.55mm, Y: 45.15mm
ROI Center Offset	X: 1.14µm, Y: 0.00µm
Reflector	AF405-49027 AF532-49014
Beam Splitter	412 550
Filter Excitation Wavelength	372 - 408 515 - 545
Filter Emission Wavelength	417 - 445 555 - 595
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @25% X-Cite Xylis Lamp Intensity @8.0%
Exposure Time	700ms 80ms
Excitation Wavelength	401 553
Emission Wavelength	422 568
Effective NA	1.25
Imaging Device	Axiocam 807 mono
Camera Adapter	1X
Depth of Focus	0.82µm 1.10µm
Binning Mode	2,2
Apotome SIM Image Processing	Display Mode "Optical Sectioning" - enabled correction selected.
3D Maximum Projection	Precise
Appearance	Threshold 10% (555), Threshold 10% (405), Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 100%, Tone Mapping was NOT Enabled
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)
Clipping Image Process	Clipping planes are introduced to pronounce regions of special interest.

FIELD	SOURCE-DOCUMENTED VALUE
X - Y	Active was selected, Textured opaque was chosen to remove the obstruction from the image above the cut plane, and a violet outline of the clipped perimeter was shown; Clip Back was Selected, and Clip Front was NOT selected.
Position of Clip	-26%
Clip X	0°
Clip Y	0°

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Identifyn® SRM Alexa Fluor™ 555 Affinity Purified Goat Anti-Rabbit, IgG (H + L)

Cat ID# - SA 000035

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher, Fish Serum Blocker cat#37527X3. The first primary antibody, Nuclear Pore Complex Monoclonal Antibody (39C7)

, was incubated overnight at 4°C at a dilution of 1:100. Cells were washed 3X in PBS. Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody followed this, Actin Recombinant Rabbit Monoclonal Antibody (JF47-01), which was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 555 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, AxioCam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 78 Slices (7.7µm)

Channels = 2

Scaling (Per Pixel) = 0.143µm X 0.143µm X 0.100µm

Image Size (Pixels) = 887 X 963

Image Size (Scaled) = 126.71µm X 137.57µm

Bit Depth = 16 Bit

Image Center Position = X: 91.55mm, Y: 45.15mm

Stage Position = 91.55mm, Y: 45.15mm

ROI Center Offset = X: 1.14µm, Y: 0.00µm

405 -

Reflector = AF405-49027
 Beam Splitter = 412
 Filter Excitation Wavelength = 372 - 408
 Filter Emission Wavelength = 417 - 445
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @25%
 Exposure Time = 700ms
 Excitation Wavelength = 401
 Emission Wavelength = 422
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 0.82µm
 Binning Mode = 2,2

555 -

Reflector = AF532-49014
 Beam Splitter = 550
 Filter Excitation Wavelength = 515 - 545
 Filter Emission Wavelength = 555 - 595
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @8.0%
 Exposure Time = 80ms
 Excitation Wavelength = 553
 Emission Wavelength = 568
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 1.10µm
 Binning Mode = 2,2

Post Processing - SIM Processing

Apotome SIM Image Processing = Display Mode "Optical Sectioning" - enabled correction selected.
 Normalization, Phase Corrector, Fourier Filter, and Deconvolution were NOT used.

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 10% (555), Threshold 10% (405), Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 100%, Tone Mapping was NOT Enabled
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Clipping Image Process = Clipping planes are introduced to pronounce regions of special interest.

X - Y = Active was selected, Textured opaque was chosen to remove the obstruction from the image above the cut plane, and a violet outline of the clipped perimeter was shown; Clip Back was Selected, and Clip Front was NOT selected.

Position of Clip = -26%

Clip X = 0°

Clip Y = 0°

Image Corrections

Identifyn® SRM Alexa Fluor™ 405 was pseudo-colored in Zen Software from dark blue, the default color, to white to provide visual contrast.


Image by J.K. Bennett

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000011
Stage	Widefield SIM - Apotome
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	43
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	BCFC2E571B9F40F73538702EDA4C86753258CA12F75B688240D56A33F4F92478
Method PDF SHA-256	46DEB090EE17F809B378B49D7F7CBD052F47875A4CEE7366D9936596614CB85A

ORIGINAL IMAGE EVIDENCE / CONFOCAL



EVIDENCE TIER	Confocal
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 750
METADATA VALUES	37

STRUCTURED ACQUISITION METADATA / CONFOCAL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	44 Slices (3.01µm)
Channels	1
Scaling (Per Pixel)	0.071µm X 0.071µm x 0.070µm
Image size (Pixels)	1330 X 1554
Image Size (Scaled)	93.90µm X 109.72µm
Bit Depth	16 Bit
Image Center Position	X: 74.30mm, Y: 51.45mm
Stage Position	X: 74.30mm, Y: 51.45mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Pinhole	1.00 AU / 36µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	405nm @0.9%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	0.90
Rotation	0°
Sampling	1.00
Pixel Time	0.33µs
Frame Time	3.95s
LSM Scan Speed	9
Scan Direction	Bidirectional
Line Step	1
Excitation Wavelength	401
Emission Wavelength	422
Effective NA	1.4
Detection Wavelength	401-422
Imaging Device	GaAsP-Pmt1
Detector Type	GaAsP-PMT
Detector Gain	750 V
Detector Offset	0

FIELD	SOURCE-DOCUMENTED VALUE
Detector Digital Gain	1.0
3D Maximum Projection	Precise
Appearance	Threshold 1% (405), Ramp 0% (405), Maximum 100%
Background	Black
Light	Brightness 80%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 59%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 44 Slices (3.01µm)

Channels = 1

Scaling (Per Pixel) = 0.071µm X 0.071µm x 0.070µm

Image size (Pixels) = 1330 X 1554

Image Size (Scaled) = 93.90µm X 109.72µm

Bit Depth = 16 Bit

Image Center Position = X: 74.30mm, Y: 51.45mm

Stage Position = X: 74.30mm, Y: 51.45mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

405 -

Reflector = None
 Contrast Method - Fluorescence
 Pinhole = 1.00 AU / 36µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 405nm @0.9%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 0.90
 Rotation = 0°
 Sampling = 1.00
 Pixel Time = 0.33µs
 Frame Time = 3.95s
 LSM Scan Speed = 9
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging - 4
 Excitation Wavelength = 401
 Emission Wavelength = 422
 Effective NA = 1.4
 Detection Wavelength = 401-422
 Imaging Device = GaAsP-Pmt1
 Detector Type = GaAsP-PMT
 Detector Gain = 750 V
 Detector Offset = 0
 Detector Digital Gain = 1.0

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 1% (405), Ramp 0% (405), Maximum 100%
 Background = Black
 Light = Brightness 80%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 59%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

Identifyn® SRM Alexa Fluor™ 405 was pseudo-colored in Zen Software from dark blue, the default color, to light blue to provide visual contrast.

Image by P. Raut

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000011
Stage	Confocal

SOURCE PROVENANCE

Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	37
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	77E3824FB39766221D834747009F9FD67DA8D2232FC3A3C2568831747286E237
Method PDF SHA-256	2DCB2712DE22C103F69BB0E1C482AB5F26A7F3A071D4DCED402C710D43666DCF

ORIGINAL IMAGE EVIDENCE / AIRYSCAN



EVIDENCE TIER	Airyscan
INSTRUMENT	ZEISS LSM 900
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488
METADATA VALUES	38

STRUCTURED ACQUISITION METADATA / AIRYSCAN

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Objective	with a Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	49 Slices (2.40µm)
Channels	1
Scaling (Per Pixel)	0.03530 µm X 0.03530 µm X 0.05000 µm
Image size (Pixels)	2389 X 1747
Image Size (Scaled)	84.33µm X 61.67µm
Bit Depth	16 Bit
Image Center Position	X: 72.22mm, Y: 51.58mm
Stage Position	X: 72.22mm, Y: 51.58mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflector	None
Pinhole	5.00 AU / 183µm
LMS MBS	MBS 405 + 488 + 561 + 640(T10/R90)
Laser Wavelength	405nm @1.0%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	1.0
Rotation	130.27°
Sampling	2.00
Pixel Time	0.73µs
Frame Time	14.22s
LSM Scan Speed	6
Scan Direction	Bidirectional
Line Step	1
Excitation Wavelength	401
Emission Wavelength	422
Effective NA	1.4
Detection Wavelength	400-471
Imaging Device	Airyscan
Detector Type	GaAsp-PMT
Detector Gain	800 V
Detector Offset	0

FIELD	SOURCE-DOCUMENTED VALUE
Detector Digital Gain	1.0
Super Resolution	8.7 (3D, Auto)
3D Maximum Projection	Precise
Appearance	Threshold 3% (405), Ramp 5% (405), Maximum 100%
Background	Black
Light	Brightness 84%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 49 Slices (2.40µm)

Channels = 1

Scaling (Per Pixel) = 35.30nm x 35.30nm x 50.00nm

Image size (Pixels) = 2389 X 1747

Image Size (Scaled) = 84.33µm X 61.67µm

Bit Depth = 16 Bit

Image Center Position = X: 72.22mm, Y: 51.58mm

Stage Position = X: 72.22mm, Y: 51.58mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

405 -

Reflector = None
 Contrast Method - Fluorescence
 Pinhole = 5.00 AU / 183µm
 LMS MBS = MBS 405 + 488 + 561 + 640(T10/R90)
 Laser Wavelength = 405nm @1.0%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.0
 Rotation = 130.27°
 Sampling = 2.00
 Pixel Time = 0.73µs
 Frame Time = 14.22s
 LSM Scan Speed = 6
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging -
 Excitation Wavelength = 401
 Emission Wavelength = 422
 Effective NA = 1.4
 Detection Wavelength = 400-471
 Imaging Device = Airyscan
 Detector Type = GaAsP-PMT
 Detector Gain = 800 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 Super Resolution: 8.7 (3D, Auto)

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 3% (405), Ramp 5% (405), Maximum 100%
 Background = Black
 Light = Brightness 84%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

Identifyn® SRM Alexa Fluor™ 405 was pseudo-colored in Zen Software from dark blue, the default color, to light blue to provide visual contrast.

Image by P. Raut

Copyright 2026 GMD12 LLC.


SOURCE PROVENANCE

Catalog	SA-000011
---------	-----------

SOURCE PROVENANCE

Stage	Airyscan
Instrument	ZEISS LSM 900
Microscope configuration	Zeiss LSM 900 Confocal, Axio Observer.Z1/7, using with a Plan-Apochromat 63X/1.40 oil DIC M27 was used with the following parameters for each dye:
Structured source metadata values	38
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	51489CC3194C011A54B4C596A3FF2128153B28D28FE449FA46BE90460C98F8A0
Method PDF SHA-256	88AED0459515834D70401D5DF983BFFC719364F0124159C11D47FE4E1A98C6B0

ORIGINAL IMAGE EVIDENCE / SIM



EVIDENCE TIER	SIM
INSTRUMENT	ZEISS Elyra
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488
METADATA VALUES	42

STRUCTURED ACQUISITION METADATA / SIM

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	94 Slices (4.65µm)
Channels	1
Scaling (Per Pixel)	0.02111 µm X 0.02111 µm X 0.05000 µm
Image Size (Pixels)	5056 X 5056
Image Size (Scaled)	106.75µm X 106.75µm
Bit Depth	16 Bit
Image Center Position	X: 60.49mm, Y: 34.67mm
Stage Position	X: 60.49mm, Y: 34.67mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflectors	SR HE LBF 405/488/561/642
Beam Splitter	405, 488, 561, 642
Filter Excitation Wavelength	200 - 405, 200 - 488, 200 - 561, 200 - 642
Filter Emission Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Contrast Method	Fluorescence
Illumination Wavelength	405
Channel Name	T2-Axiocam 820#2-SR
Excitation Wavelength	405
Emission Wavelength	445
Effective NA	1.4
Detection Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Imaging Device	Axiocam 820#2
Camera Adapter	1X
Exposure Time	100ms
Depth of Focus	0.69µm
Binning Mode	2,2
Laser Wavelength	405nm @3.0%
Frame Time	676.00ms
Scan Direction	Unidirectional
Result Sampling	4
Process Sampling	N/A
Algorithm	SIM

FIELD	SOURCE-DOCUMENTED VALUE
SIM Settings	Standard
Auto Sharpness	On
Sharpness	9.6872 (405)
Fitting	Fast Fit
Histogram	Scale to Min/Max
Baseline	Cut
3D Maximum Projection	Precise
Appearance	Threshold 1% (405), Ramp 0% (405), Maximum 100%
Background	Black
Light	Brightness 100%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 94 Slices (4.65µm)

Channels = 1

Scaling (Per Pixel) = 16.89nm X 16.89nm X 50.00nm

Image Size (Pixels) = 5056 X 5056

Image Size (Scaled) = 106.75µm X 106.75µm

Bit Depth = 16 Bit

Image Center Position = X: 60.49mm, Y: 34.67mm

Stage Position = X: 60.49mm, Y: 34.67mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

405 -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 405
 Channel Name = T2-Axiocam 820#2-SR
 Excitation Wavelength = 405
 Emission Wavelength = 445
 Effective NA = 1.4
 Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Imaging Device = Axiocam 820#2
 Camera Adapter = 1X
 Exposure Time = 100ms
 Depth of Focus = 0.69µm
 Binning Mode = 2,2
 Laser Wavelength = 405nm @3.0%
 Frame Time = 676.00ms
 Scan Direction = Unidirectional
 Grating 36.5µm grating

Post Processing - SIM Processing

Processing Dimension - 3D
 Result Sampling = 4
 Process Sampling = N/A
 Channels = 1
 Algorithm = SIM
 SIM Settings = Standard
 Auto Sharpness = On
 Sharpness = 9.6872 (405)
 Fitting = Fast Fit
 Histogram = Scale to Min/Max
 Baseline = Cut

3D Image Processing

3D Maximum Projection = Precise
 Appearance = Threshold 1% (405), Ramp 0% (405), Maximum 100%
 Background = Black
 Light = Brightness 100%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 10%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

None


Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000011
Stage	SIM
Instrument	ZEISS Elyra
Microscope configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Structured source metadata values	42
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	A578F9C949FFFE81E0F6A5136B709767A98A48EBC99ACC9F33616972517BC15E
Method PDF SHA-256	D75F8DE8725279963A295C3433ABB0869ADD4B33C46A3701A3D3ADEA3D9C0873

ORIGINAL IMAGE EVIDENCE / SIM²



EVIDENCE TIER	SIM ²
INSTRUMENT	ZEISS Elyra
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488
METADATA VALUES	45

STRUCTURED ACQUISITION METADATA / SIM²

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	62 Slices (3.05µm)
Channels	1
Scaling (Per Pixel)	0.02111 µm X 0.02111 µm X 0.05000 µm
Image Size (Pixels)	3525 X 3946
Image Size (Scaled)	74.42µm X 83.81µm
Bit Depth	16 Bit
Image Center Position	X: 61.22mm, Y: 35.26mm
Stage Position	X: 61.22mm, Y: 35.26mm
ROI Center Offset	X: 0.00µm, Y: 0.00µm
Reflectors	SR HE LBF 405/488/561/642
Beam Splitter	405, 488, 561, 642
Filter Excitation Wavelength	200 - 405, 200 - 488, 200 - 561, 200 - 642
Filter Emission Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Contrast Method	Fluorescence
Illumination Wavelength	405
Channel Name	T2-Axiocam 820#2-SR
Excitation Wavelength	405
Emission Wavelength	445
Effective NA	1.4
Detection Wavelength	419 - 470, 503 - 546, 576 - 617, 657 - 800
Imaging Device	Axiocam 820#2
Camera Adapter	1X
Exposure Time	100ms
Depth of Focus	0.69µm
Binning Mode	2,2
Laser Wavelength	405nm @3.0%
Frame Time	676.00ms
Scan Direction	Unidirectional
Result Sampling	4
Process Sampling	4
Algorithm	SIM2

FIELD	SOURCE-DOCUMENTED VALUE
SIM Settings	Standard
Auto Sharpness	On
Sharpness	9.0203 (405)
Order Combination	Weiner Filter
Regularization	None
Iterations	3
Filter	No Filter
Fitting	Fast Fit
Histogram	Scale to Min/Max
3D Volume Projection	Precise
Appearance	Threshold 1% (405), , Ramp 98% (405, Maximum 100%
Background	Black
Light	Brightness 95%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Microscope

Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:

Z Stack - (3D)

Z-Stack = 62 Slices (3.05µm)

Channels = 1

Scaling (Per Pixel) = 21.11nm X 21.11nm X 50.00nm

Image Size (Pixels) = 3525 X 3946

Image Size (Scaled) = 74.42µm X 83.81µm

Bit Depth = 16 Bit

Image Center Position = X: 61.22mm, Y: 35.26mm

Stage Position = X: 61.22mm, Y: 35.26mm

ROI Center Offset = X: 0.00µm, Y: 0.00µm

405 -

Reflectors = SR HE LBF 405/488/561/642
 Beam Splitter = 405, 488, 561, 642
 Filter Excitation Wavelength = 200 - 405, 200 - 488, 200 - 561, 200 - 642
 Filter Emission Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Contrast Method = Fluorescence
 Illumination Wavelength = 405
 Channel Name = T2-Axiocam 820#2-SR
 Excitation Wavelength = 405
 Emission Wavelength = 445
 Effective NA = 1.4
 Detection Wavelength = 419 - 470, 503 - 546, 576 - 617, 657 - 800
 Imaging Device = Axiocam 820#2
 Camera Adapter = 1X
 Exposure Time = 100ms
 Depth of Focus = 0.69µm
 Binning Mode = 2,2
 Laser Wavelength = 405nm @3.0%
 Frame Time = 676.00ms
 Scan Direction = Unidirectional
 Grating 36.5µm grating

Post Processing - SIM2 Processing

Processing Dimension - 3D
 Result Sampling = 4
 Process Sampling = 4
 Channels = 1
 Algorithm = SIM2
 SIM Settings = Standard
 Auto Sharpness = On
 Sharpness = 9.0203 (405)
 Order Combination = Weiner Filter
 Regularization = None
 Iterations = 3
 Filter = No Filter
 Fitting = Fast Fit
 Histogram = Scale to Min/Max

3D Image Processing

3D Volume Projection = Precise
 Appearance = Threshold 1% (405), , Ramp 98% (405, Maximum 100%
 Background = Black
 Light = Brightness 95%, Azimuth 80°, Elongation -130°, Enabled Directional Lighting, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Image Corrections

Identifyn® SRM Alexa Fluor™ 405 was pseudo-colored in Zen Software from dark blue, the default color, to light blue to provide visual contrast.


Image by P. Raut

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000011
Stage	SIM²
Instrument	ZEISS Elyra
Microscope configuration	Zeiss Elyra 5 Lattice SIM, AxioObserver.Z1/ 7 using Plan-Apochromat 63X/1.40 oil DIC M27, 1X Tubelens, was used with the following parameters for each dye:
Structured source metadata values	45
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	B1F8ACE8C801D6CFF109D30CB5CC1079067126C1EB9B2B98FE8F7909FCA7052A
Method PDF SHA-256	F260AC075AAD5D71F40BB4CAAF680ADAEA4E3B27A67A31F3912A16EFE5649069

ORIGINAL IMAGE EVIDENCE / CONTROL



EVIDENCE TIER	Control
INSTRUMENT	ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 647
METADATA VALUES	29

STRUCTURED ACQUISITION METADATA / CONTROL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Channels	2
Scaling (Per Pixel)	0.143µm X 0.143µm
Image Size (Pixels)	1192 X 947
Image Size (Scaled)	170.29µm X 135.29µm
Bit Depth	14 Bit
Image Center Position	X: 99.81mm, Y: 54.82mm
ROI Center Offset	X: 99.81µm, Y: 54.82µm
Reflector	AF405-49027 50 Cy5
Beam Splitter	412 660
Filter Excitation Wavelength	372 - 408 625 - 655
Filter Emission Wavelength	417 - 445 665 - 715
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @25%
Exposure Time	650ms 200ms
Excitation Wavelength	401 653
Emission Wavelength	422 668
Effective NA	1.25
Imaging Device	Axiocam 807 mono
Camera Adapter	1X
Depth of Focus	0.82µm 1.30µm
Binning Mode	2,2

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® SRM Alexa Fluor™ 405 Affinity Purified Rabbit Anti-Mouse F(ab')₂ (H + L)

Cat ID# - SA 000011

Identifyn® SRM Alexa Fluor™ 647 Affinity Purified Goat Anti-Mouse, IgG (H + L), (Subclasses 1+2a+2b+3), Fcy Fragment Specific

Cat ID# - SA 000056

Hela Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. In the absence of a primary antibody, the Identifyn® 405 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS.

To demonstrate the presence of cells in the sample, minus DAPI (which would overlap the 405), a second primary antibody followed this, Alpha Tubulin Monoclonal Antibody (DM1A), which was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® 647 secondary antibody was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using Prolong™ Diamond Antifade Mountant cat#36961.

Note that the lamp intensity and exposure times were set to the 647 control samples.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D) -

Channels = 2

Scaling (Per Pixel) = 0.143µm X 0.143µm

Image Size (Pixels) = 1192 X 947

Image Size (Scaled) = 170.29µm X 135.29µm

Bit Depth = 14 Bit

Image Center Position = X: 99.81mm, Y: 54.82mm

ROI Center Offset = X: 99.81µm, Y: 54.82µm

405-

Reflector = AF405-49027
 Beam Splitter = 412
 Filter Excitation Wavelength = 372 - 408
 Filter Emission Wavelength = 417 - 445
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @25%
 Exposure Time = 650ms
 Excitation Wavelength = 401
 Emission Wavelength = 422
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 0.82µm
 Binning Mode = 2,2

647 -

Reflector = 50 Cy5
 Beam Splitter = 660
 Filter Excitation Wavelength = 625 - 655
 Filter Emission Wavelength = 665 - 715
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @25%
 Exposure Time = 200ms
 Excitation Wavelength = 653
 Emission Wavelength = 668
 Effective NA = 1.25
 Imaging Device = Axiocam 807 mono
 Camera Adapter = 1X
 Depth of Focus = 1.30µm
 Binning Mode = 2,2

Post Processing -

None

Image Corrections

None

Control Summary -

In the absence of a primary antibody, Identifyn Standards for Optical Microscopy Methods™ were applied using standard staining methodology and image acquisition settings using the Zeiss Axio Imager.M1 Widefield. With no image correction, the image demonstrates that the secondary antibody, minus a primary target, had no off-target binding and no significant presence of free dye within the sample.

The data suggests the secondary antibody is highly specific and free of unconjugated (free) dye.

Image by J.K. Bennett

Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000011
Stage	Control
Instrument	ZEISS Axio Imager.M1; ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Axiocam 807, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	29
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	5258559E7435F4B97D6B623EAA21A562ED124F157EC01D06B3B9CF6445C986EB
Method PDF SHA-256	88F0C7CFE7B5A210E9AE820CEF47CCD01BBBCD5C9CFE4A33F4A66E764D10BC14