

IDENTIFYN RESOLUTION STEPDOWN

SECONDARY ANTIBODY MICROSCOPY EVIDENCE ASSESSMENT

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)

WESTERN BLOT -> WIDEFIELD -> WIDEFIELD SIM / APOTOME -> CONFOCAL

4


ORDERED EVIDENCE TIERS

1

SUPPORTING CONTROL

PENDING

SCIENTIST REVIEW



REPRESENTATIVE HIGHEST-AVAILABLE VIEW / CONFOCAL

SA-000106 | PRODUCT-LEVEL EVIDENCE DOSSIER

SOURCE-GROUNDED MICROSCOPY + BIOLOGICAL REVIEW / IDENTIFYN EVIDENCE PRESERVED

CRO CAPABILITY DEMONSTRATED

Identifyn performed the experimental and imaging work documented below. BENNETT converts that operating history into a governed, source-bounded scientific record.

- Source-documented sample preparation and antibody/reagent workflow
- Western blot imaging
- Widefield fluorescence microscopy
- Widefield structured illumination / Apotome optical sectioning
- Confocal microscopy
- Preserved control experiment
- Treatment-response experimental workflow
- Z-stack / 3D acquisition
- Multi-channel imaging

SCIENTIFIC QUESTION

What source-bounded experimental and microscopy evidence is preserved for Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L), and what is the strongest interpretation allowed by the available modalities and controls?

EVIDENCE AVAILABLE

Western blot -> Widefield -> Widefield SIM — Apotome -> Confocal -> Control

BENNETT ASSESSMENT

The preserved SA-000106 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal. The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

BENNETT / MICROSCOPY AND BIOLOGICAL EVIDENCE REVIEW

SOURCE-GROUNDED RETROSPECTIVE REVIEW / PENDING SCIENTIST REVIEW

SOURCE STATE	SOURCE HASH VERIFIED / EVENT RECORDED
BENNETT EVIDENCE REVIEW	COMPLETED / SOURCE-BOUNDED
NATIVE IMAGE REPROCESSING	NOT PERFORMED
LITERATURE SOURCES	INDEPENDENTLY VERIFIED
SCIENTIST REVIEW	PENDING
PUBLICATION	NOT AUTHORIZED

OVERALL CALL / NIR-BOUNDED PARTIAL STEPDOWN

The preserved SA-000106 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal. The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

SETTINGS ASSESSMENT

The record preserves 4 ordered evidence tier(s). Structured source metadata values are reported by stage and remain tied to the original method objects.

BIOLOGICAL SUPPORT

The record supports retrospective, source-bounded microscopy review. Target identity, specificity, treatment effect, binding, interaction and mechanism are not inferred from presentation images alone.

CONTROL ASSESSMENT

A true preserved control record is present and appears last.

CLAIM CEILING

This draft supports historical capability documentation, source registration, modality ordering, settings review and bounded interpretation. It does not independently validate antibody specificity, quantify a population response, establish binding, interaction, causation or mechanism, or authorize publication.

PROCESSING DISCLOSURE

Retrospective BENNETT architectural evaluation of preserved Identifyn source evidence. BENNETT did not perform native-file reprocessing, new deterministic measurement or the historical image post-processing represented in this record.

REAGENTS AND ACQUIRED CHANNELS

Preparation reagents are reported separately from channels actually documented in the acquisition settings. Fluorophores mentioned in staining are not automatically treated as acquired channels.

EVIDENCE OBJECT	REAGENTS / FLUOROPHORES PRESENT	CHANNELS ACTUALLY ACQUIRED
Western blot	790	Not deterministically stated in the acquisition block
Widefield	405/DAPI; 532; 555; 790	3; AF750-49007; AF532-49014; 49 DAPI; 672 - 748; 515 - 545; 335 - 383; 765 - 855
Widefield SIM — Apotome	405/DAPI; 532; 555; 790	3; AF750-49007; AF532-49014; 672 - 748; 515 - 545; 765 - 855; 555 - 595; 752
Confocal	405/DAPI; 488; 532; 790	3; None; 730 nm @0.8%; 488 nm @0.7%; 405 nm @0.2%; 789; 528; 353
Control	405/DAPI; 790	2; None; 730 nm @0.8%; 405 nm @0.2%; 789; 353; 794; 465

MODALITY ASSESSMENT / PART 1 OF 5

Western blot / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: Azure Sapphire FL. Objective: not explicitly stated. Platform: Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s). Structured source metadata values: 11.

VISIBLE OBSERVATION

A preserved presentation image is available for Western blot. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Western blot, documents platform context as Azure Sapphire FL, and records detected dye/channel descriptors as 790.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Band position and displayed intensity do not independently establish analytical specificity, target identity, linearity, or treatment effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Western blot evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 2 OF 5

Widefield / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:. Acquisition: Channels: 3; Scaling (Per Pixel): 0.073 μm X 0.073 μm ; Image size (Pixels): 2256 X 2219; Image Size (Pixels): 2256 X 2219; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Hamamatsu ORCA-Quest. Timing: Exposure Time: 250 ms; 50 ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @25.00%; X-Cite Xylis Lamp Intensity @5.00%. Optical/scan: Effective NA: 1.25. Structured source metadata values: 40.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 532; 555; 790.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

A presentation image supports gross localization and source-bounded co-occurrence, not direct binding, molecular colocalization, or native quantitative measurement.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 3 OF 5

Widefield SIM — Apotome / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS Axio Observer.Z1/7 Apotome. Objective: EC Plan-Neofluar 63X/1.25 oil M27. Platform: Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:. Acquisition: Z-Stack: 73 Slices (7.2 μm); Channels: 3; Scaling (Per Pixel): 0.073 μm X 0.073 μm X 0.100 μm ; Image size (Pixels): 2236 X 2035; Image Size (Pixels): 2236 X 2035; Bit Depth: 16 Bit. Detector/camera: Imaging Device: Hamamatsu ORCA-Quest. Timing: Exposure Time: 200ms; 100 ms. Illumination: Light source: X-Cite Xylis Lamp Intensity @20.00%; X-Cite Xylis Lamp Intensity @10.00%. Optical/scan: Effective NA: 1.25. Processing: Projection: View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black. Structured source metadata values: 46.

VISIBLE OBSERVATION

A preserved presentation image is available for Widefield SIM — Apotome. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Widefield SIM — Apotome, documents platform context as ZEISS Axio Observer.Z1/7 Apotome, and records detected dye/channel descriptors as 405/DAPI; 532; 555; 790.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Improved optical sectioning does not by itself establish reconstructed super-resolution or molecular interaction.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Widefield SIM — Apotome evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 4 OF 5

Confocal / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 980. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:. Acquisition: Z-Stack: 201 Slices (6 μm); Channels: 3; Scaling (Per Pixel): 0.059 μm X 0.059 μm X 0.030 μm ; Image size (Pixels): 930 X 930; Image Size (Pixels): 930 X 930; Bit Depth: 16 Bit. Detector/camera: Imaging Device: NIR Ch2; GaAsP-Pmt3; Detector Type: GaAs-PMT; GaAsP-PMT. Timing: Pixel Time: 0.36 μs ; Frame Time: 3.08 s. Illumination: Laser Wavelength: 730 nm @0.8%; 488 nm @0.7%. Optical/scan: Pinhole: 1.00 AU / 79 μm ; 1.00 AU / 53 μm ; Scan Zoom: 2.4; LSM Scan Speed: 12; Scan Direction: Bidirectional; Averaging: 8; Effective NA: 1.4. Processing: Projection: View Angle 35°, Scale Z 109%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected); Background: Black; LSM Plus Mode: 7.4 (3D, Auto); 7.1 (3D, Auto). Structured source metadata values: 64.

VISIBLE OBSERVATION

A preserved presentation image is available for Confocal. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Confocal, documents platform context as ZEISS LSM 980, and records detected dye/channel descriptors as 405/DAPI; 488; 532; 790.

LITERATURE CORRELATION

Consistent with peer-reviewed microscopy and antibody-validation guidance that presentation evidence must remain bounded by modality, controls and source documentation.

INTERPRETIVE LIMIT

Optical sectioning and source-bounded 3D presentation do not establish binding, mechanism, or population-level effect.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Confocal evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

MODALITY ASSESSMENT / PART 5 OF 5

Control / SOURCE EVIDENCE PRESERVED

DOCUMENTED SETTINGS

Platform: ZEISS LSM 980. Objective: Plan-Apochromat 63X/1.40 oil DIC M27. Platform: Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:. Acquisition: Channels: 2; Scaling (Per Pixel): 0.059 μm X 0.059 μm ; Image size (Pixels): 1612 X 1612; Image Size (Pixels): 1612 X 1612; Bit Depth: 16 Bit. Detector/camera: Imaging Device: NIR Ch2; GaAsP-Pmt3; Detector Type: GaAs-PMT; GaAsP-PMT. Timing: Pixel Time: 0.24 μs ; Frame Time: 6.04 s. Illumination: Laser Wavelength: 730 nm @0.8%; 405 nm @0.2%. Optical/scan: Pinhole: 1.00 AU / 79 μm ; 1.00 AU / 44 μm ; Scan Zoom: 1.2; 1.4; LSM Scan Speed: 11; Scan Direction: Bidirectional; Averaging: 8; Effective NA: 1.4. Processing: LSM Plus Mode: 6.9 (3D, Auto); 6.0 (3D, Auto). Structured source metadata values: 47.

VISIBLE OBSERVATION

A preserved presentation image is available for Control. BENNETT records the visible source presentation without creating a new native measurement.

SOURCE-REPORTED RESULT

The source identifies this object as Control, documents platform context as ZEISS LSM 980, and records detected dye/channel descriptors as 405/DAPI; 790.

LITERATURE CORRELATION

Consistent with peer-reviewed antibody-control guidance that a no-primary or related control bounds background but is not a complete specificity system.

INTERPRETIVE LIMIT

A control bounds the recorded comparison only; it does not establish universal specificity, zero background, zero cross-reactivity, or absence of free dye.

CLAIM CEILING

This tier supports a source-bounded statement that the historical Control evidence and documented settings are preserved for review; stronger biological or quantitative conclusions require compatible native data, controls and scientist review.

CROSS-MODALITY SYNTHESIS

The preserved SA-000106 record contains the applicable NIR sequence: Western blot; Widefield; Widefield SIM — Apotome; Confocal. The historical boundary reflects the validated application available for the specific conjugate, deployed platform and workflow, not a universal scientific impossibility.

Different acquisitions are not treated as a quantitative intensity ladder. Any continuity statement remains qualitative and source bounded.

BENNETT LITERATURE CORRELATION — SOURCES INDEPENDENTLY VERIFIED

BENNETT uses independently verified primary scientific sources to test whether the interpretation is scientifically consistent and properly bounded. Literature correlation does not retroactively validate the Identifyn dataset or replace scientist review.

- Antibody and microscopy evidence require orthogonal controls and modality-appropriate interpretation; presentation images alone do not establish analytical specificity.
- Advanced-resolution presentation depends on acquisition, reconstruction or localization quality control before numerical resolution, molecular dimensions or cluster identity can be claimed.

REFERENCES

1. Uhlen M, et al. A proposal for validation of antibodies. Nat Methods. 2016;13:823-827. doi:10.1038/nmeth.3995.
2. Gustafsson MGL. Surpassing the lateral resolution limit by a factor of two using structured illumination microscopy. J Microsc. 2000;198:82-87. doi:10.1046/j.1365-2818.2000.00710.x.
3. Sage D, et al. Quantitative evaluation of software packages for single-molecule localization microscopy. Nat Methods. 2015;12:717-724. doi:10.1038/nmeth.3442.

RELEASE BOUNDARY

This assessment remains a draft pending scientist review. No publication is authorized. Any future native-file analysis must enter BENNETT as a new evidence snapshot with routed engine authority, native-data QA, methods, limitations, and an independently reviewed claim ceiling.

SOURCE RECONCILIATION / QA / PROVENANCE

SOURCE STATUS - PRESERVED / AUTO-RECONCILED

Historical Identifyn source material was subjected to BENNETT retrospective QA. Original source state and correction history remain preserved in provenance.

HIGH CONFIDENCE - SCIENTIST-AUTHORIZED CORRECTION

Cross-tier / treatment concentration: 200 μ M; 2 μ M -> Etoposide = 2 μ M. The governing scientist identified every 200 μ M occurrence as a technician transcription error. The canonical historical concentration is 2 μ M. Supporting evidence: Scientist-authorized correction supplied for the Identifyn historical archive; original technician text remains preserved unchanged. Authority: Dr. Brian T. Bennett. Preservation: Original source text and hashes remain unchanged; only the derivative canonical field is corrected.

UNRESOLVED - PRESERVED AS CONFLICT

Cross-tier / treatment duration: {"Western blot": ["20 hours"], "Widefield": ["24 hours"], "Widefield SIM — Apotome": ["24 hours"], "Confocal": ["24 hours"]} -> Preserved as modality-specific historical duration values. Genuinely different treatment durations are present across evidence stages, and no source or scientist correction resolves them. Supporting evidence: Only duration values from explicit etoposide-treatment sentences were classified; concentration, wavelength, resolution and pixel-size values were excluded.

BENNETT EVIDENCE REVIEW COMPLETE

INTERPRETATION TO EVIDENCE

BENNETT has completed the bounded retrospective assessment.
The following pages switch ownership to the preserved Identifyn record.



IDENTIFYN PRESERVED SOURCE EVIDENCE

ORIGINAL RECORD BEGINS

Images, source methods, acquisition settings and provenance follow in the locked
Stepdown order. Scientific wording is preserved; missing modalities are not invented.
A preserved control follows all available Stepdown tiers as supporting evidence.

PRODUCT-LEVEL STEPDOWN MAP


This dossier preserves the complete available evidence progression for SA-000106. Each tier remains tied to its original image, source method, instrument context and cryptographic identity. A preserved control is retained as supporting evidence and appears after every resolution tier.

CANONICAL STEPDOWN	EVIDENCE STAGE	INSTRUMENT	STATUS
1	Western blot	Azure Sapphire FL	PRESERVED
2	Widefield	ZEISS Axio Observer.Z1/7 Apotome	PRESERVED
3	Widefield SIM - Apotome	ZEISS Axio Observer.Z1/7 Apotome	PRESERVED
4	Confocal	ZEISS LSM 980	PRESERVED

INTERPRETIVE BOUNDARY

Ordering evidence by resolution tier does not make the modalities interchangeable. Each stage retains its acquisition environment and scientific limitations. This pilot organizes the record; it does not synthesize new biological claims.

ORIGINAL IMAGE EVIDENCE / WESTERN BLOT



EVIDENCE TIER	Western blot
INSTRUMENT	Azure Sapphire FL
CELL MODEL	HeLa
DETECTED DYES	790
METADATA VALUES	11

STRUCTURED ACQUISITION METADATA / WESTERN BLOT

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Scanner	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Detection mode	Fluorescence
Laser	730nm
Emission Filter	829±62nm
Detector	PMT
Intensity	3
Pixel size	100 µm
Scan Speed	High
Quality	High
Focus	Membrane (0 µm in Z-axis)
Image	K. Small

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Secondary Antibody

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)

Cat ID# - SA 000106

HeLa cells were treated with 200 μM etoposide for 20 hours to induce DNA damage. Cells were then homogenized using an ultrasonication probe and RIPA lysis buffer, and then centrifuged to pellet the cell debris. The lysate supernatant (100 μg) was mixed with NuPAGE™ LDS Sample Buffer (4X), heated at 80 °C for 10 min, and subjected to electrophoresis using NuPAGE™ 4 to 12% Bis-Tris, 1.0-1.5 mm, Mini Protein Gels in a Mini Gel Tank. The Thermo Fisher's PageRuler™ Prestained Protein Ladder, 10 to 180 kDa, was used as a molecular weight marker.

Post electrophoresis, the proteins were transferred to a polyvinylidene difluoride (PVDF) membrane using the iBlot™ 3 Starter Kit PVDF. The transfer membrane was blocked with Blocker™ Casein for 1 hour and then washed five times with PBS. The blocked membrane was incubated with Actin Recombinant Rabbit Monoclonal Antibody (JF47-01) for 2 hours, followed by five washes in 1X PBS. Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L) was incubated for 1 hour and washed 5X with PBS. The membrane was dried for 10 minutes at room temperature before imaging.

Scanner = Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)

Scan Settings

Detection mode = Fluorescence

Laser = 730nm

Emission Filter = 829±62nm

Detector = PMT

Intensity = 3

Pixel size = 100 μm

Scan Speed = High

Quality = High

Focus = Membrane (0 μm in Z-axis)

Post Processing

None


Image: K. Small

© Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000106
Stage	Western blot
Instrument	Azure Sapphire FL
Microscope configuration	Azure Sapphire FL with Customizable Laser and Filter Modules, Enabling UV- NIR and Z-Plane(s)
Structured source metadata values	11
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	57BC9B069D68D9910CB6C421AEA3FD0F3449D81626C29B93F8FE1613A9056F26
Method PDF SHA-256	809465A61BF1A63042D8B7C619029D855FC15FCCABE4FE42D13601E8AF4238B3

ORIGINAL IMAGE EVIDENCE / WIDEFIELD



Identifyn™ SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)
Cat ID# - SA 000106

Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N)
Cat ID# - PA-RR 000007

Identifyn™ SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
Cat ID# - SA 000026

EVIDENCE TIER	Widefield
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 532; 555; 790
METADATA VALUES	40

STRUCTURED ACQUISITION METADATA / WIDEFIELD

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Channels	3
Scaling (Per Pixel)	0.073 μ m X 0.073 μ m
Image size (Pixels)	2256 X 2219
Image Size (Scaled)	164.72 μ m X 162.02 μ m
Bit Depth	16 Bit
Image Center Position	X: 0.00 μ m, Y: 0.00 μ m
ROI Center Offset	X: 0.00 μ m, Y: 0.00 μ m
Reflector	AF750-49007 AF532-49014 49 DAPI
Beam Splitter	760 550 395
Filter Excitation Wavelength	672 - 748 515 - 545 335 - 383
Filter Emission Wavelength	765 - 855 555 - 595 420 - 470
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @25.00% X-Cite Xylis Lamp Intensity @5.00% X-Cite Xylis Lamp Intensity @10.00%
Exposure Time	250 ms 50 ms 25ms
Excitation Wavelength	752 528 353
Emission Wavelength	779 553 465
Effective NA	1.25
Imaging Device	Hamamatsu ORCA-Quest
Camera Adapter	1X
Depth of Focus	1.51 μ m 1.07 μ m 0.90 μ m
Binning Mode	1,1
Profile Definition	Stroke Thickness 1, Normal View, Grid Distance 1 Profile View = X and Y axes were set to Auto - X (Min 0.0 and Max 132.6), Y (Min 311 and Max 6825)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N)

Cat ID# - PA-RR 000007

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)

Cat ID# - SA 000106

Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L)

Cat ID# - SA 000026

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2µM Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L) was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was then incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L) was incubated at room temperature for 1 hour at a dilution of 1:100. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI, cat#P36962.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D)

Channels = 3

Scaling (Per Pixel) = 0.073 µm X 0.073 µm

Image size (Pixels) = 2256 X 2219

Image Size (Scaled) = 164.72 µm X 162.02 µm

Bit Depth = 16 Bit

Image Center Position = X: 0.00 µm, Y: 0.00 µm

ROI Center Offset = X: 0.00 µm, Y: 0.00 µm

790 -

Reflector = AF750-49007
 Beam Splitter = 760
 Filter Excitation Wavelength = 672 - 748
 Filter Emission Wavelength = 765 - 855
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @25.00%
 Exposure Time = 250 ms
 Excitation Wavelength = 752
 Emission Wavelength = 779
 Effective NA = 1.25
 Imaging Device = Hamamatsu ORCA-Quest
 Camera Adapter = 1X
 Depth of Focus = 1.51 μ m
 Binning Mode = 1,1

532 -

Reflector = AF532-49014
 Beam Splitter = 550
 Filter Excitation Wavelength = 515 - 545
 Filter Emission Wavelength = 555 - 595
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @5.00%
 Exposure Time = 50 ms
 Excitation Wavelength = 528
 Emission Wavelength = 553
 Effective NA = 1.25
 Imaging Device = Hamamatsu ORCA-Quest
 Camera Adapter = 1X
 Depth of Focus = 1.07 μ m
 Binning Mode = 1,1

DAPI -

Reflector = 49 DAPI
 Beam Splitter = 395
 Filter Excitation Wavelength = 335 - 383
 Filter Emission Wavelength = 420 - 470
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @10.00%
 Exposure Time = 25ms
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.25
 Imaging Device = Hamamatsu ORCA-Quest
 Camera Adapter = 1X
 Depth of Focus = 0.90 μ m
 Binning Mode = 1,1

Profile View Display - Panels A & B

Profile Definition = Stroke Thickness 1, Normal View, Grid Distance 1 Profile View = X and Y axes were set to Auto - X (Min 0.0 and Max 132.6), Y (Min 311 and Max 6825)

Split View - Panel C

The top-left panel features Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N) visualized with Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L), the top-right panel features alpha tubulin visualized with Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L), the bottom-left panel shows DAPI nuclear staining, and the bottom-right panel is the merged image of all channels.

Image Correction

None

Image by E.F. Simon

Image Analysis by B.T. Bennett


Data Presentation by P.D. Amistoso

© Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000106
Stage	Widefield
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	40
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	F6E7F829A7683B4AA5974BCDE304F6C7D4B7916C602057C7C1AEE4FA21EF3183
Method PDF SHA-256	869F9AB3D60D384DE87049431E07C2F12C6D84B4EDA9D9E6EE5F4F8017A98E2C

ORIGINAL IMAGE EVIDENCE / WIDEFIELD SIM - APOTOME



EVIDENCE TIER	Widefield SIM - Apotome
INSTRUMENT	ZEISS Axio Observer.Z1/7 Apotome
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 532; 555; 790
METADATA VALUES	46

STRUCTURED ACQUISITION METADATA / WIDEFIELD SIM - APOTOME

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Objective	EC Plan-Neofluar 63X/1.25 oil M27
Z-Stack	73 Slices (7.2 μm)
Channels	3
Scaling (Per Pixel)	0.073 μm X 0.073 μm X 0.100 μm
Image Size (Pixels)	2236 X 2035
Image Size (Scaled)	163.26 μm X 148.59 μm
Bit Depth	16 Bit
Image Center Position	X: 0.00 mm, Y: 0.00 mm
ROI Center Offset	X: 0.00 μm , Y: 0.00 μm
Reflector	AF750-49007 AF532-49014
Beam Splitter	760 550
Filter Excitation Wavelength	672 - 748 515 - 545
Filter Emission Wavelength	765 - 855 555 - 595
Contrast Method	Fluorescence
Light source	X-Cite Xylis Lamp Intensity @20.00% X-Cite Xylis Lamp Intensity @10.00%
Exposure Time	200ms 100 ms
Excitation Wavelength	752 528
Emission Wavelength	779 553
Effective NA	1.25
Imaging Device	Hamamatsu ORCA-Quest
Camera Adapter	1X
Depth of Focus	1.51 μm 1.07 μm
Binning Mode	1,1
Apotome SIM Image Processing	Display Mod" "Optical Sectioning" - enabled correction selected.
3D Maximum Projection	Precise
Appearance	Threshold 13% all channels, Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 100%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)
Clipping Image Process	Clipping planes are introduced to pronounce regions of special interest.
X-Z	Activate was selected, textured opaque was chosen, and a green outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.

FIELD	SOURCE-DOCUMENTED VALUE
Position of Clip	-9% -7%
Clip X	60° 0°
Clip Z	0°
X-Y	Activate was selected, textured opaque was chosen, and a blue outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.
Clip Y	0°

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N)

Cat ID# - PA-RR 000007

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)

Cat ID# - SA 000106

Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L)

Cat ID# - SA 000026

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2µM Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human BRCA1 Recombinant Monoclonal, IgG - (BLR023N), was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L) was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was then incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L) was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:

Single Plane (2D) - Z-plane 40 of 73 total - Panel C

Z Stack - (3D) - Panels A, B, & C

Z-Stack = 73 Slices (7.2 µm)

Channels = 3

Scaling (Per Pixel) = 0.073 µm X 0.073 µm X 0.100 µm

Image Size (Pixels) = 2236 X 2035

Image Size (Scaled) = 163.26 µm X 148.59 µm

Bit Depth = 16 Bit

Image Center Position = X: 0.00 mm, Y: 0.00 mm

ROI Center Offset = X: 0.00 µm, Y: 0.00 µm

790 -

Reflector = AF750-49007
 Beam Splitter = 760
 Filter Excitation Wavelength = 672 - 748
 Filter Emission Wavelength = 765 - 855
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @20.00%
 Exposure Time = 200ms
 Excitation Wavelength = 752
 Emission Wavelength = 779
 Effective NA = 1.25
 Imaging Device = Hamamatsu ORCA-Quest
 Camera Adapter = 1X
 Depth of Focus = 1.51 μ m
 Binning Mode = 1,1

532 -

Reflector = AF532-49014
 Beam Splitter = 550
 Filter Excitation Wavelength = 515 - 545
 Filter Emission Wavelength = 555 - 595
 Contrast Method = Fluorescence
 Light source = X-Cite Xylis Lamp Intensity @10.00%
 Exposure Time = 100 ms
 Excitation Wavelength = 528
 Emission Wavelength = 553
 Effective NA = 1.25
 Imaging Device = Hamamatsu ORCA-Quest
 Camera Adapter = 1X
 Depth of Focus = 1.07 μ m
 Binning Mode = 1,1

Post Processing - SIM Processing

Apotome SIM Image Processing = Display Mod" "Optical Sectioning" - enabled correction selected.
 Normalization, Phase Corrector, Fourier Filter, and Deconvolution were NOT used.

Histogram View - Panel A

Histogram Definition - Normal, Histo Table was Frequency, Bin Count 256, Bin Size 256, Lower Threshold 0, Upper Threshold 65535, Showing Threshold, and deploying Logarithmic Binning

Histogram View - X and Y axis auto, Channel Transparency 26%, Statistic Table, Frequency Table, Internal Measurements, and Image were selected.

3D Image Processing - Panel B

3D Maximum Projection = Precise

Appearance = Threshold 13% all channels, Ramp 0% all channels, Maximum 100% all channels

Background = Black

Light = Brightness 100%, Enabled Tone Mapping

Projection = View Angle 35°, Scale Z 100%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Clipping Image Process = Clipping planes are introduced to pronounce regions of special interest.

X-Z = Activate was selected, textured opaque was chosen, and a green outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.

Position of Clip = -9%

Clip X = 60°

Clip Z = 0°

X-Y = Activate was selected, textured opaque was chosen, and a blue outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.

Position of Clip = -7%

Clip X = 0°

Clip Y = 0°

Summary

Panel A

Histogram - Measures the frequency and intensity of the signal within the sample. In this analysis, we assess the entire image located in the upper right corner for both the 790 and 532 channels to provide comparative intensities. We focus on the Rabbit anti-Human BRCA1 Recombinant Monoclonal IgG (BLR023N) monoclonal staining, which is counterstained with the Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit F(ab')₂ (H + L). Additionally, we examine Alpha Tubulin, which is counterstained with the Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse IgG (H + L). Notably, there is a strong signal for the BRCA1/790 immunostaining, observed at the approximate center z-plane, which is shown as red (background or behind using a transparency for the 532-signal shown in the forefront as white). This indicates a diffuse pan-nuclear protein, especially when compared to the robust amounts of Alpha Tubulin, represented by a white overlay using the Identifyn® SRM Alexa Fluor™ 790, which demonstrates high brightness.

Image Correction

None

Image by E.F. Simon

Image Analysis by B.T. Bennett


Data Presentation by P.D. Amistoso

© Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000106
Stage	Widefield SIM - Apotome
Instrument	ZEISS Axio Observer.Z1/7 Apotome
Microscope configuration	Zeiss Axio Observer.Z1/7 Apotome SIM, using an EC Plan-Neofluar 63X/1.25 oil M27, Hamamatsu ORCA-Quest, and X-Cite Xylis Lamp was used with the following parameters for each dye:
Structured source metadata values	46
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	4B8BD3B6A7ECE94B55B66786606A9CC45AD7FEF71EEB5C7A9DCEEDBD42BB305D
Method PDF SHA-256	A722DDEAD76E9AB9B75621E451445B8DA965DE198FF187CC49658860DCE0B22B

ORIGINAL IMAGE EVIDENCE / CONFOCAL



EVIDENCE TIER	Confocal
INSTRUMENT	ZEISS LSM 980
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 488; 532; 790
METADATA VALUES	64

STRUCTURED ACQUISITION METADATA / CONFOCAL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Z-Stack	201 Slices (6 µm)
Channels	3
Scaling (Per Pixel)	0.059 µm X 0.059 µm X 0.030 µm
Image Size (Pixels)	930 X 930
Image Size (Scaled)	54.71 µm X 54.71 µm
Bit Depth	16 Bit
Image Center Position	X: 89.08 mm, Y: 49.76 mm
ROI Center Offset	X: 0.00 µm, Y: 0.00 µm
Reflector	None
Contrast Method	Fluorescence
Pinhole	1.00 AU / 79 µm 1.00 AU / 53 µm 1.00 AU / 44 µm
LSM MBS	MBS 488/561, 405/730
LSM SBS	SBS LP 740 Mirror
Laser Wavelength	730 nm @0.8% 488 nm @0.7% 405 nm @0.2%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	2.4
Rotation	0°
Sampling	1.20
Pixel Time	0.36 µs
Frame Time	3.08 s
LSM Scan Speed	12
Scan Direction	Bidirectional
Line Step	1
Averaging	8
Excitation Wavelength	789 528 353
Emission Wavelength	794 553 465
Effective NA	1.4
Detection Wavelength	755-900 530-580 430-499
Imaging Device	NIR Ch2 GaAsP-Pmt3
Detector Type	GaAs-PMT GaAsP-PMT

FIELD	SOURCE-DOCUMENTED VALUE
Detector Gain	700 V 500 V
Detector Offset	0
Detector Digital Gain	1.0
LSM Plus Mode	7.4 (3D, Auto) 7.1 (3D, Auto) 6.3 (3D, Auto)
3D Maximum Projection	Precise
Appearance	Threshold 2% all channels, Ramp 0% all channels, Maximum 100% all channels
Background	Black
Light	Brightness 111%, Enabled Tone Mapping
Projection	View Angle 35°, Scale Z 109%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)
Clipping Image Process	Clipping planes are introduced to pronounce regions of special interest.
X-Z	Activate was selected, textured opaque was chosen, and a green outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.
Position of Clip	20% 22%
Clip X	0°
Clip Z	0°
Y-Z	Activate was selected, textured opaque was chosen, and a red outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.
Clip Y	0°

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Primary Antibody

Rabbit anti-Human Ku70 Polyclonal, IgG
Cat ID# - PA-RP 000006

Identifyn® Secondary Antibodies

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)
 Cat ID# - SA 000106

Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L)
 Cat ID# - SA 000026

HeLa cells were grown on acid-etched #1.5H cover glass to ~50% confluency. DNA damage was induced via treatment with 2μM Etoposide-spiked complete growth media for 24 hours, followed by -20°C methanol fixation and blocked with Thermo Fisher Fish Serum Blocker cat#37527X3. The primary antibody, Rabbit anti-Human Ku70 Polyclonal, IgG, was incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS, and Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L) was incubated at room temperature for 1 hour at a dilution of 1:1000.

The second primary antibody, Alpha Tubulin Monoclonal Antibody (DM1A), was then incubated for 1 hour at room temperature at a dilution of 1:250. Cells were washed 3X in PBS. Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L) was incubated at room temperature for 1 hour at a dilution of 1:100. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:

Single Plane (2D) - Z-plane 101 of 201 total - Panel C

Z Stack - (3D) - Panels A, B, & C

Z-Stack = 201 Slices (6 μm)

Channels = 3

Scaling (Per Pixel) = 0.059 μm X 0.059 μm X 0.030 μm

Image Size (Pixels) = 930 X 930

Image Size (Scaled) = 54.71 μm X 54.71 μm

Bit Depth = 16 Bit

Image Center Position = X: 89.08 mm, Y: 49.76 mm

Image Center Position = X: 89.08 mm, Y: 49.76 mm

ROI Center Offset = X: 0.00 μm, Y: 0.00 μm

790 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00 AU / 79 μ m
 LSM MBS = MBS 488/561, 405/730
 LSM SBS = SBS LP 740
 Laser Wavelength = 730 nm @0.8%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 2.4
 Rotation = 0°
 Sampling = 1.20
 Pixel Time = 0.36 μ s
 Frame Time = 3.08 s
 LSM Scan Speed = 12
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 8
 Excitation Wavelength = 789
 Emission Wavelength = 794
 Effective NA = 1.4
 Detection Wavelength = 755-900
 Imaging Device = NIR Ch2
 Detector Type = GaAs-PMT
 Detector Gain = 700 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 LSM Plus Mode = 7.4 (3D, Auto)

532 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00 AU / 53 μ m
 LSM MBS = MBS 488/561, 405/730
 LSM SBS = Mirror
 Laser Wavelength = 488 nm @0.7%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 2.4
 Rotation = 0°
 Sampling = 1.20
 Pixel Time = 0.36 μ s
 Frame Time = 3.08 s
 LSM Scan Speed = 12
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 8
 Excitation Wavelength = 528
 Emission Wavelength = 553
 Effective NA = 1.4
 Detection Wavelength = 530-580
 Imaging Device = GaAsP-Pmt3
 Detector Type = GaAsP-PMT
 Detector Gain = 500 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 LSM Plus Mode = 7.1 (3D, Auto)

DAPI -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00 AU / 44 µm
 LSM MBS = MBS 488/561, 405/730
 LSM SBS = Mirror
 Laser Wavelength = 405 nm @0.2%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 2.4
 Rotation = 0°
 Sampling = 1.20
 Pixel Time = 0.36 µs
 Frame Time = 3.08 s
 LSM Scan Speed = 12
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 8
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Detection Wavelength = 430-499
 Imaging Device = GaAsP-Pmt3
 Detector Type = GaAsP-PMT
 Detector Gain = 500 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 LSM Plus Mode = 6.3 (3D, Auto)

3D Image Processing - Panel A

3D Maximum Projection = Precise
 Appearance = Threshold 2% all channels, Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 111%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 109%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

3D Image Processing - Panel B

3D Maximum Projection = Precise
 Appearance = Threshold 2% all channels, Ramp 0% all channels, Maximum 100% all channels
 Background = Black
 Light = Brightness 111%, Enabled Tone Mapping
 Projection = View Angle 35°, Scale Z 109%, Reverse Z-Axis NOT Selected, Stereo Anaglyph NOT Selected, Camera Separation 5% (deselected), Parallax Shift 0% (deselected)

Clipping Image Process = Clipping planes are introduced to pronounce regions of special interest.

X-Z = Activate was selected, textured opaque was chosen, and a green outline of the clipped perimeter was shown;
 Clip Back was selected, and Clip Front was NOT selected.

Position of Clip = 20%

Clip X = 0°

Clip Z = 0°

Y-Z = Activate was selected, textured opaque was chosen, and a red outline of the clipped perimeter was shown; Clip Back was selected, and Clip Front was NOT selected.

Position of Clip = 22%

Clip Y = 0°

Clip Z = 0°

Split View - Panel C

The first panel on the left features Rabbit anti-Human Ku70 Polyclonal, IgG visualized with Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L), the second panel features alpha tubulin visualized with Identifyn® SRM Alexa Fluor™ 532 Affinity Purified Donkey Anti-Mouse, IgG (H + L), the third panel shows DAPI nuclear staining, and the last panel on the right shows the merged image of all channels. This image was taken at Z-Position 101 of 201.

Image Corrections

Panel B - DAPI was pseudo-colored in Zen Software from blue, the default color, to dark grey to enhance visual contrast.

Image by J.K. Bennett

Image Analysis by B.T. Bennett


Data Presentation by P.D. Amistoso

© Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000106
Stage	Confocal
Instrument	ZEISS LSM 980
Microscope configuration	Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:
Structured source metadata values	64
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	3CC2C7CF476F2780DFA1741CB6CA657ABA177D9F204B9B50E19D3C4C266CC9B3
Method PDF SHA-256	BB632F7B0CA6187A4DB6E1E0994ECE520ED7542E49C2A769D40F296A8825EE53

ORIGINAL IMAGE EVIDENCE / CONTROL



EVIDENCE TIER	Control
INSTRUMENT	ZEISS LSM 980
CELL MODEL	HeLa
DETECTED DYES	405/DAPI; 790
METADATA VALUES	47

STRUCTURED ACQUISITION METADATA / CONTROL

The values below were deterministically extracted from the preserved Identifyn method PDF text. They remain source-documented metadata, not new BENNETT measurements.

FIELD	SOURCE-DOCUMENTED VALUE
Platform / configuration	Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:
Objective	Plan-Apochromat 63X/1.40 oil DIC M27
Channels	2
Scaling (Per Pixel)	0.059 μm X 0.059 μm
Image size (Pixels)	1612 X 1612
Image Size (Scaled)	94.79 μm X 94.79 μm
Bit Depth	16 Bit
Image Center Position	X: 93.65 μm , Y: 48.24 μm
Stage Position	X: 93.65 μm , Y: 48.24 μm
ROI Center Offset	X: 0.00 μm , Y: 0.00 μm
Reflector	None
Contrast Method	Fluorescence
Pinhole	1.00 AU / 79 μm 1.00 AU / 44 μm
LSM MBS	Plate, MBS 405/730
LSM SBS	SBS LP 740 Mirror
Laser Wavelength	730 nm @0.8% 405 nm @0.2%
Laser Blanking	Enabled
Scan Mode	Frame
Scan Zoom	1.2 1.4
Rotation	0°
Sampling	1.20
Pixel Time	0.24 μs
Frame Time	6.04 s
LSM Scan Speed	11
Scan Direction	Bidirectional
Line Step	1
Averaging	8
Excitation Wavelength	789 353
Emission Wavelength	794 465
Effective NA	1.4
Detection Wavelength	755-900 430-490
Imaging Device	NIR Ch2 GaAsP-Pmt3
Detector Type	GaAs-PMT GaAsP-PMT

FIELD	SOURCE-DOCUMENTED VALUE
Detector Gain	700 V 500 V
Detector Offset	0
Detector Digital Gain	1.0
LSM Plus Mode	6.9 (3D, Auto) 6.0 (3D, Auto)
Profile Definition	Stroke Thickness 1.00, Normal View, Grid Distance 1 Profile View = X and Y axes were set to Auto - X (Min 0.0 and Max 87.3), Y (Min 0 and Max 20668)

SOURCE METHOD

The text below is carried from the preserved Identifyn source record. Typography and pagination have been normalized for this preservation dossier; scientific wording has not been editorially rewritten.

Identifyn Standards for Optical Microscopy Methods™

Identifyn® Secondary Antibody

Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L)

Cat ID# - SA 000106

HeLa Cells were grown on acid-etched #1.5H cover glass to ~50% confluency, followed by -20°C methanol fixation and blocked with Thermo Fisher, Fish Serum Blocker cat#37527X3. In the absence of a primary antibody, Identifyn® SRM Alexa Fluor™ 790 Affinity Purified Goat Anti-Rabbit, F(ab')₂ (H + L) was incubated at room temperature for 1 hour at a dilution of 1:1000. Cells were washed 5X in PBS and mounted using ProLong™ Diamond Antifade Mountant with DAPI cat#P36962.

Microscope

Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:

Single Plane (2D)

Channels = 2

Scaling (Per Pixel) = 0.059 µm X 0.059 µm

Image size (Pixels) = 1612 X 1612

Image Size (Scaled) = 94.79 µm X 94.79 µm

Bit Depth = 16 Bit

Image Center Position = X: 93.65 µm, Y: 48.24 µm

Stage Position = X: 93.65 µm, Y: 48.24 µm

ROI Center Offset = X: 0.00 µm, Y: 0.00 µm

790 -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00 AU / 79 μ m
 LSM MBS = Plate, MBS 405/730
 LSM SBS = SBS LP 740
 Laser Wavelength = 730 nm @0.8%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.2
 Rotation = 0°
 Sampling = 1.20
 Pixel Time = 0.24 μ s
 Frame Time = 6.04 s
 LSM Scan Speed = 11
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 8
 Excitation Wavelength = 789
 Emission Wavelength = 794
 Effective NA = 1.4
 Detection Wavelength = 755-900
 Imaging Device = NIR Ch2
 Detector Type = GaAs-PMT
 Detector Gain = 700 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 LSM Plus Mode = 6.9 (3D, Auto)

DAPI -

Reflector = None
 Contrast Method = Fluorescence
 Pinhole = 1.00 AU / 44 μ m
 LSM MBS = Plate, MBS 405/730
 LSM SBS = Mirror
 Laser Wavelength = 405 nm @0.2%
 Laser Blanking = Enabled
 Scan Mode = Frame
 Scan Zoom = 1.4
 Rotation = 0°
 Sampling = 1.20
 Pixel Time = 0.24 μ s
 Frame Time = 6.04 s
 LSM Scan Speed = 11
 Scan Direction = Bidirectional
 Line Step = 1
 Averaging = 8
 Excitation Wavelength = 353
 Emission Wavelength = 465
 Effective NA = 1.4
 Detection Wavelength = 430-490
 Imaging Device = GaAsP-Pmt3
 Detector Type = GaAsP-PMT
 Detector Gain = 500 V
 Detector Offset = 0
 Detector Digital Gain = 1.0
 LSM Plus Mode = 6.0 (3D, Auto)

Profile View Display - Panels A & B

Profile Definition = Stroke Thickness 1.00, Normal View, Grid Distance 1 Profile View = X and Y axes were set to Auto
 - X (Min 0.0 and Max 87.3), Y (Min 0 and Max 20668)

Image Correction

None

Image by E.F. Simon
 Image Analysis by B.T. Bennett
 Data Presentation by P.D. Amistoso

© Copyright 2026 GMD12 LLC.

SOURCE PROVENANCE

Catalog	SA-000106
Stage	Control

SOURCE PROVENANCE

Instrument	ZEISS LSM 980
Microscope configuration	Zeiss LSM 980 Confocal with NIR Detection, Axio Observer.Z1/7, using a Plan-Apochromat 63X/1.40 oil DIC M27, was used with the following parameters for each dye:
Structured source metadata values	47
Metadata source	PRESERVED_SOURCE_METHOD_PDF_TEXT
Cell model	HeLa
Image SHA-256	5F54648824AEC65D7FE02C1F527735B17A9E24CC10C18D2528893BC00672AE45
Method PDF SHA-256	5F45B4CF17308577F5E0D35D8E3EC73DA30BE06E009273F6FDB5198FCDE5EB21